



Datasheet for ABIN1000268

## Peroxide Assay Kit



[Go to Product page](#)

1 Image

16 Publications

### Overview

Quantity:	250 tests
Target:	Peroxide
Application:	Biochemical Assay (BCA)

### Product Details

Sample Type:	Serum, Plasma (citrate), Urine, Cell Lysate, Cell Culture Supernatant
Specificity:	0.4 $\mu$ M
Characteristics:	<p>Sensitive and accurate. Enhanced color intensity using sorbitol. Detection range 0.2 <math>\mu</math>M (7 ng/mL) to 30 <math>\mu</math>M (1,020 ng/mL) H<sub>2</sub>O<sub>2</sub> in 96-well plate assay.</p> <p>Simple and high-throughput. The procedure involves addition of a single detection reagent and incubation for 30 min. Can be readily automated as a high-throughput assay in 96-well plates for thousands of samples per day.</p>
Components:	Reagent A: 1 mL. Reagent B: 50 mL. Standard: 100 $\mu$ L 3% stabilized H <sub>2</sub> O <sub>2</sub> .
Material not included:	Pipetting devices and accessories, 96-well plates and plate reader.

### Target Details

Target:	Peroxide
Background:	<p>Quantitative determination of peroxide by colorimetric (585nm) method.</p> <p>Procedure: 30 min.</p>

Peroxide (e.g. hydrogen peroxide H<sub>2</sub>O<sub>2</sub>) is one of the key reactive oxygen species formed under oxidative stress conditions. High levels of peroxide formation have been linked to pathological

## Target Details

---

conditions such as ageing, asthma, diabetes, atherosclerosis, cataract, inflammatory arthritis and neurodegenerative diseases. Simple, direct and automation-ready procedures for quantitative determination of peroxide find wide applications in research and drug discovery. This peroxide assay kit is designed to measure peroxide concentration in biological samples without any pretreatment. The improved method utilizes the chromogenic Fe<sup>3+</sup>-xylenol orange reaction, in which a purple complex is formed when Fe<sup>2+</sup> provided in the reagent is oxidized to Fe<sup>3+</sup> by peroxides present in the sample. The intensity of the color, measured at 540-610nm, is an accurate measure of the peroxide level in the sample. The optimized formulation substantially reduces interference by substances in the raw samples.

## Application Details

---

Application Notes:	Direct Assays: H <sub>2</sub> O <sub>2</sub> in biological samples (e.g. serum, citrate-plasma, urine, cell lysate, culture medium). Pharmacology: effects of drugs on peroxide metabolism.
Protocol:	Procedure using 96-well plate: <ol style="list-style-type: none"><li>Standards. Prepare fresh standards on the day of assay. Pipette 5 µL 3% H<sub>2</sub>O<sub>2</sub> and mix well with 495 µL H<sub>2</sub>O in a 1.5-mL Eppendorf tube. Mix 5 µL of this solution with 1465 µL H<sub>2</sub>O. The final H<sub>2</sub>O<sub>2</sub> concentration is 30 µM (labeled Premix). Dilute standard as shown in the Table.</li><li>Transfer 40 µL diluted standards and each sample into separate wells of a clear flat-bottom 96-well plate. Add 200 µL Detection Reagent to all standards and samples.</li><li>Incubate 30 min at room temperature and read optical density at 540-610nm (peak absorbance at 585nm). Note: if in rare cases, precipitation occurs after adding the Detection Reagent to a sample, transfer the whole reaction mixture of this sample well into a 1.5-mL Eppendorf tube and centrifuge 2 min at 14,000 rpm. Carefully remove 200 µL supernatant into a clean well and read OD. Multiply the OD reading by 1.2 to account for the volume change.</li></ol>
Reagent Preparation:	Equilibrate to room temperature before assay. Prepare enough Detection Reagent by mixing 1 volume of Reagent A with 100 volumes Reagent B.
Sample Preparation:	Several chemicals are known to interfere and should be avoided in sample preparation. These include ascorbic acid, EDTA, heparin, DMSO (>0.02%), NP-40 (>0.6%), SDS (>0.12%), Tris (>8mM) and ethanol (>0.4%). Samples can be analyzed immediately after collection, or stored in aliquots at -20 °C. Avoid repeated freeze-thaw cycles.
Calculation of Results:	Subtract blank OD (water, #8) from the standard OD values and plot the OD against H <sub>2</sub> O <sub>2</sub> concentrations. Subtract blank OD from Sample OD. Determine the sample peroxide content from the standard curve.

## Application Details

Conversions: 1  $\mu\text{M}$  H<sub>2</sub>O<sub>2</sub> equals 34 ng/mL or 34 ppb.

Restrictions: For Research Use only

## Handling


Storage: 4 °C

## Publications

Product cited in: Li, Uddayasankar, He, Wang, Qin: "Fast, Sensitive, and Quantitative Point-of-Care Platform for the Assessment of Drugs of Abuse in Urine, Serum, and Whole Blood." in: **Analytical chemistry**, Vol. 89, Issue 16, pp. 8273-8281, (2019) ([PubMed](#)).

There are more publications referencing this product on: [Product page](#)

## Images



### Biochemical Assay

Image 1.