antibodies - online.com









Image



| () | 1/0 | r\ /1 | 014 | |
|-----|-----|-------|-----|---|
| () | ve | I V I | -v | V |
| | | | | |

| Quantity: | 100 tests |
|--------------|--|
| Target: | ITGA3 |
| Reactivity: | Human |
| Host: | Please inquire |
| Clonality: | Monoclonal |
| Conjugate: | This ITGA3 antibody is conjugated to PE |
| Application: | Flow Cytometry (FACS), Immunofluorescence (IF) |

Product Details

| Clone: | VJ1-6 | |
|------------------|--|--|
| Isotype: | lgG1 | |
| Characteristics: | CD49c, a 150kDa cell surface antigen which is also known as the alpha-3 integrin and as VLA-3. CD49c is expressed by most cell types including T and B lymphocytes | |

Target Details

| Target: | ITGA3 |
|-------------------|--|
| Alternative Name: | CD49c (ITGA3 Products) |
| Background: | The antibody recognizes the VLA-a3 chain of the VLA-3 complex and is reactive with epithelial tissue and basal renal distal tubules. |
| Pathways: | CXCR4-mediated Signaling Events, Integrin Complex |

Application Details

| Application Notes: | It is recommended for use in flow cytometry. This reagent is effective for direct immunofluorescence staining of human tissue for flow cytometric analysis using 20 μ l/10^6 cells. | |
|---------------------|--|--|
| Comment: | R-Phycoerythrin (Europa Bioproducts, Ely, Cambridge). | |
| Sample Preparation: | 1. Transfer 100 μl of anticoagulated (EDTA) blood to a 12 x 75 mm polystyrene test tube (10 cells). 2. Add 20 μl of CD49c PE and mix gently with a vortex mixer. The 20 μl is a guideline of the optimal volume should be determined by the individual laboratory. 3. The recommended negative control is a non-reactive PE-conjugated antibody of the same isotype. 4. Incubate in the dark at room temperature at 4°C for 30 minutes or at room temperature (20-25 °C) for 1 minutes. 5. Add 1,5 ml of Lysing Solution to each sample and mix gently with a vortex mixer Incubate for 10 minutes at room temperature in the dark. 6. Centrifuge at 1000 x g for 5 minutes. Gently aspirate the supernatant and discard it leaving approximately 50 μl of fluid. Add 2 ml 0.01 mol/l PBS (It better that it containing 2% bovine serum albumin) and resusper the cells by using a vortex mixer.8. Centrifuge at 1000 x g for 5 minutes. Gently aspirate the supernatant and discard it leaving approximately 50 μl of fluid. 9. Resuspend pellet in an appropriate fluid for flow cytometry, e.g. 0.3 ml PBS. The PBS should contain 1% paraformaldehyde (fixative) if samples are not analysed the same day. 10. Analyse on a flow cytometer or store at 2-8 °C in the dark until analysis. Samples can be run up to 24 hours aft lysis. | |
| Restrictions: | For Research Use only | |
| Handling | | |
| Format: | Liquid | |
| Buffer: | The conjugate is provided in liquid form in buffer containing 1% bovine serum albumin (BSA) and 0,09% Sodium azide, pH 7.2. | |
| Preservative: | Sodium azide | |
| Precaution of Use: | 1. The device is not intended for clinical use including diagnosis, prognosis, and monitoring of a disease state, and it must not be used in conjunction with patient records or treatment. 2. This product contains Sodium azide (NaN3), a chemical highly toxic in pure form. At product concentrations, though not classified as hazardous, Sodium azide may react with lead and copper plumbing to form highly explosive build-ups of metal azides. Upon disposal, flush with | |

large volumes of water to prevent metal azide build-up in plumbing. 3. As with any product

derived from biological sources, proper handling procedures should be used.

Storage:

4°C

Images




Image 1.