# antibodies -online.com







# **CCL17 ELISA Kit**



Go to	Prod	uct	page

$\sim$					
( )	VE	۲۱	/1	$\triangle$	Λ

Quantity:	96 tests
Target:	CCL17
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	31.2-2000 pg/mL
Minimum Detection Limit:	31.2 pg/mL
Application:	ELISA
Product Details	
Purpose:	For quantitative detection of CCL17 in human serum, plasma, body fluids, tissue lysates or cell
	culture supernatants.
Sample Type:	Cell Culture Supernatant, Plasma, Serum, Tissue Lysate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Sensitivity:	< 1 pg/mL
Components:	1. One 96-well plate pre-coated with anti-Human CCL17 antibody 2. Lyophilized Human CCL17
	standards: 2 tubes (10ng / tube) 3. Sample / Standard diluent buffer: 30ml 4. Biotin conjugated
	anti-Human CCL17 antibody (Concentrated): 130 µl.
Material not included:	1. 37 °C incubator 2. Microplate reader (wavelength: 450nm) 3. Precise pipette and disposable
	pipette tips 4. Automated plate washer 5. ELISA shaker 6. 1.5ml of Eppendorf tubes 7. Plate
	cover 8. Absorbent filter papers 9. Plastic or glass container with volume of above 1L

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn | International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com | Page 1/3 | Product datasheet for ABIN1112572 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

## **Target Details**

l arget Details	
Target:	CCL17
Alternative Name:	CCL17 (CCL17 Products)
Background:	Chemokine (C-C motif) ligand 17 (CCL17), also known as thymus and activation regulated chemokine (TARC), is a small cytokine belonging to the CC chemokine family. It is expressed constitutively in thymus, but only transiently in phytohemagglutinin-stimulated peripheral blood mononuclear cells. This chemokine specifically binds and induces chemotaxis in T cells and elicits its effects by interacting with the chemokine receptor CCR4. The gene for CCL17 maps to 16q13.
Application Details	
Comment:	This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti-CCL17 polyclonal antibody was pre-coated onto 96-well plates. And the biotin conjugated anti-CCL17 polyclonal antibody was used as detection antibodies. The standards test samples and biotin conjugated detection antibody were added - the wells subsequently and wash with wash buffer Avidin-Biotin-Peroxidase Complex was added and unbound conjugates were washed away with wash buffer. TMB substrates were used - visualize HRP enzymatic reaction. TMB was catalyzed by HRP - produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional - the CCL17 amount of sample captured in plate. Read the O.D. absorbance at 450 nm in a microplate reader and then the concentration of CCL17 can be calculated.
Plate:	Pre-coated
Reagent Preparation:	1. Before the experiment, centrifuge each kit component for several minutes to bring down all reagents to the bottom of tubes. 2. It is recommend to measure each standard and sample in duplicate. 3. Do NOT let the plate completely dry at any time! Since the dry condition can inactivate the biological material on the plate. 4. Do not reuse pipette tips and tubes to avoid cross contamination. 5. Do not use the expired components and the components from different batches. 6. To avoid the marginal effect of plate incubation for temperature differences (the marginal wells always get stronger reaction), it is recommend to equilibrate the ABC working solution and TMB substrate for at least 30 min at room temperature (37°C) before adding to wells. The TMB substrate (Kit Component 8) is colorless and transparent before use, if not, please contact us for replacement.
Sample Preparation:	Preparation of sample and reagents 1. Sample Isolate the test samples soon after collecting, then, analyze immediately (within 2 hours). Or aliquot and store at -20 °C for long term. Avoid

### **Application Details**

multiple freeze-thaw cycles.

Body fluids, tissue lysates or Cell culture supernatants: Centrifuge to remove precipitate, analyze immediately or aliquot and store at -20  $^{\circ}$ C .

Serum: Coagulate the serum at room temperature (about 4 hours). Centrifuge at approximately  $1000 \times g$  for 15 min. Analyze the serum immediately or aliquot and store at -20 °C .

Plasma: Collect plasma with EDTA as the anticoagulant. Centrifuge for 15 min at 1000 x g within 30 min of collection. Analyze immediately or aliquot and store frozen at -20 °C. Citrate and heparin can not be used as anticoagulant here. Note: 1. Coagulate blood samples completely, then, centrifuge, and avoid hemolysis and particle. 2. NaN3 can not be used as test sample preservative, since it is the inhibitor for HRP.

Restrictions:

For Research Use only

### Handling

Preservative:

Sodium azide, Thimerosal (Merthiolate)