# antibodies -online.com





## Datasheet for ABIN1112607

## **CX3CL1 ELISA Kit**



### Overview

Quantity:	96 tests
Target:	CX3CL1
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	156-10000 pg/mL
Minimum Detection Limit:	156 pg/mL
Application:	ELISA
Product Details	
Purpose:	For quantitative detection of Fractalkine in human serum, plasma, body fluids, tissue lysates or cell culture supernatants.
Sample Type:	Cell Culture Supernatant, Plasma, Serum, Tissue Lysate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Sensitivity:	< 50 pg/mL
Components:	1. One 96-well plate pre-coated with anti-human fractalkine antibody 2. Lyophilized human fractalkine standards: 2 tubes (10ng / tube) 3. Sample / Standard diluent buffer: 30ml 4. Biotin conjugated anti-human fractalkine antibody (Concentrated): 130 µl.
Material not included:	1. 37 °C incubator 2. Microplate reader (wavelength: 450nm) 3. Precise pipette and disposable pipette tips 4. Automated plate washer 5. ELISA shaker 6. 1.5ml of Eppendorf tubes 7. Plate cover 8. Absorbent filter papers 9. Plastic or glass container with volume of above 1L

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN1112607 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

### **Target Details**

l arget Details	
Target:	CX3CL1
Alternative Name:	Fractalkine / CX3CL1 (CX3CL1 Products)
Background:	Chemokine (C-X3-C motif) ligand 1 (CX3CL1), also known as fractalkine (in humans) and
	neurotactin (in mice), is a large cytokine protein of 373 amino acids, it contains multiple
	domains and is the only known member of the CX3C chemokine family. Its gene is located on
	human chromosome 16 along with some CC chemokines known as CCL17 and CCL22.
	CX3CL1 is produced as a long protein (with 373-amino acid in humans) with an extended
	mucin-like stalk and a chemokine domain on top. The mucin-like stalk permits it to bind to the
	surface of certain cells. However a soluble (90 kD) version of this chemokine has also been
	observed. CX3CL1 elicits its adhesive and migratory functions by interacting with the
	chemokine receptor CX3CR1.
Pathways:	Synaptic Membrane
Application Details	
Comment:	This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti-
	fractalkine polyclonal antibody was pre-coated onto 96-well plates. And the biotin conjugated
	anti-fractalkine polyclonal antibody was used as detection antibodies. The standards test
	samples and biotin conjugated detection antibody were added - the wells subsequently and
	wash with wash buffer. Avidin-Biotin-Peroxidase Complex was added and unbound conjugates
	were washed away with wash buffer. TMB substrates were used - visualize HRP enzymatic
	reaction. TMB was catalyzed by HRP - produce a blue color product that changed into yellow
	after adding acidic stop solution. The density of yellow is proportional - the fractalkine amount
	of sample captured in plate. Read the O.D. absorbance at 450 nm in a microplate reader and
	then the concentration of fractalkine can be calculated.
Plate:	Pre-coated
Reagent Preparation:	1. Before the experiment, centrifuge each kit component for several minutes to bring down all
	reagents to the bottom of tubes. 2. It is recommend to measure each standard and sample in
	duplicate. 3. Do NOT let the plate completely dry at any time! Since the dry condition can
	inactivate the biological material on the plate. 4. Do not reuse pipette tips and tubes to avoid
	cross contamination. 5. Do not use the expired components and the components from differe
	batches. 6. To avoid the marginal effect of plate incubation for temperature differences (the
	marginal wells always get stronger reaction), it is recommend to equilibrate the ABC working

solution and TMB substrate for at least 30 min at room temperature (37°C) before adding to

#### **Application Details**

wells. The TMB substrate (Kit Component 8) is colorless and transparent before use, if not, please contact us for replacement.

#### Sample Preparation:

Preparation of sample and reagents 1. Sample Isolate the test samples soon after collecting, then, analyze immediately (within 2 hours). Or aliquot and store at -20 °C for long term. Avoid multiple freeze-thaw cycles.

Tissue lysate, body fluids and cell culture supernatants: Centrifuge to remove precipitate, analyze immediately or aliquot and store at -20 °C.

Serum: Coagulate the serum at room temperature (about 4 hours). Centrifuge at approximately  $1000 \times g$  for 15 min. Analyze the serum immediately or aliquot and store at -20 °C .

Plasma: Collect plasma with heparin as the anticoagulant. Centrifuge for 15 min at  $1000 \times g$  within 30 min of collection. Analyze immediately or aliquot and store frozen at -20 °C. Citrate and EDTA can not be used as anticoagulant here. Note: 1. Coagulate blood samples completely, then, centrifuge, and avoid hemolysis and particle. 2. NaN3 can not be used as test sample preservative, since it is the inhibitor for HRP.

Restrictions:

For Research Use only

#### Handling

Preservative:

Sodium azide, Thimerosal (Merthiolate)