antibodies -online.com





Datasheet for ABIN1112656

IL-7 ELISA Kit



Overview

Quantity:	96 tests
Target:	IL-7 (IL7)
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	15.6-1000 pg/mL
Minimum Detection Limit:	15.6 pg/mL
Application:	ELISA

Product Details

Purpose:	For quantitative detection of IL-7 in Human serum, plasma or cell.
Sample Type:	Cell Culture Cells, Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Sensitivity:	<1 pg/mL
Components:	1. One 96-well plate pre-coated with anti-Human IL-7 antibody 2. Lyophilized Human IL-7 standards: 2 tubes (10ng / tube) 3. Sample / Standard diluent buffer: 30ml 4. Biotin conjugated anti-Human IL-7 antibody (Concentrated): 130 µl.
Material not included:	1. 37 °C incubator 2. Microplate reader (wavelength: 450nm) 3. Precise pipette and disposable pipette tips 4. Automated plate washer 5. ELISA shaker 6. 1.5ml of Eppendorf tubes 7. Plate cover 8. Absorbent filter papers 9. Plastic or glass container with volume of above 1L

Target Details

IL-7 (IL7) Target: Alternative Name: IL-7 (IL7 Products) Background: Interleukin 7(IL-7) is a hematopoietic growth factor secreted by stromal cells in the red marrow and thymus. It is also produced by keratinocytes, dendritic cells, hepatocytes, neurons, and epithelial cells. IL-7 is a cytokine important for B and T cell development. This cytokine and the hepatocyte growth factor (HGF) form a heterodimer that functions as a pre-pro-B cell growthstimulating factor. This cytokine is found to be a cofactor for V(D)J rearrangement of the T cell receptor beta (TCRß) during early T cell development. II-7 promotes hematological malignacies (acute lymphoblastic leukemia, T cell lymphoma), and the elevated levels of IL-7 have also been detected in the plasma of HIV-infected patients. **JAK-STAT Signaling** Pathways: **Application Details** Comment: This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti-IL-7 polyclonal antibody was pre-coated onto 96-well plates. And the biotin conjugated anti-IL-7 polyclonal antibody was used as detection antibodies. The standards test samples and biotin conjugated detection antibody were added - the wells subsequently and wash with wash buffer. Avidin-Biotin-Peroxidase Complex was added and unbound conjugates were washed away with wash buffer. TMB substrates were used - visualize HRP enzymatic reaction. TMB was catalyzed by HRP - produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional - the IL-7 amount of sample captured in plate. Read the O.D. absorbance at 450 nm in a microplate reader and then the concentration of IL-7 can be calculated. Plate: Pre-coated Reagent Preparation: 1. Before the experiment, centrifuge each kit component for several minutes to bring down all reagents to the bottom of tubes. 2. It is recommend to measure each standard and sample in duplicate. 3. Do NOT let the plate completely dry at any time! Since the dry condition can inactivate the biological material on the plate. 4. Do not reuse pipette tips and tubes to avoid cross contamination. 5. Do not use the expired components and the components from different batches. 6. To avoid the marginal effect of plate incubation for temperature differences (the marginal wells always get stronger reaction), it is recommend to equilibrate the ABC working

solution and TMB substrate for at least 30 min at room temperature (37°C) before adding to

wells. The TMB substrate (Kit Component 8) is colorless and transparent before use, if not,

Application Details

Preparation:

Preparation of sample and reagents 1. Sample Isolate the test samples soon after collecting, then, analyze immediately (within 2 hours). Or aliquot and store at -20 °C for long term. Avoid multiple freeze-thaw cycles.

Cell: Lyse cells thoroughly, centrifuge at 10000-14000 X g for 3-5 min, collect supernatant, analyze immediately or aliquot and store at -20 °C.

Serum: Coagulate the serum at room temperature (about 30 min). Centrifuge at approximately 2000 × g for 20 min. Analyze the serum immediately or aliquot and store at -20 °C.

Plasma: Collect plasma with citrate, heparin or EDTA as the anticoagulant. Centrifuge for 15 min at 1000 x g within 30 min of collection. Analyze immediately or aliquot and store frozen at -20 °C. Note: 1. Coagulate blood samples completely, then, centrifuge, and avoid hemolysis and particle. 2. NaN3 can not be used as test sample preservative, since it is the inhibitor for HRP.

Restrictions:

For Research Use only

Handling

Preservative: Sodium azide, Thimerosal (Merthiolate)