

Datasheet for ABIN1112776 P-Cadherin ELISA Kit



[Go to Product page](#)

Overview

| | |
|--------------------------|-------------------|
| Quantity: | 96 tests |
| Target: | P-Cadherin (CDH3) |
| Reactivity: | Human |
| Method Type: | Sandwich ELISA |
| Detection Range: | 62.5-4000 pg/mL |
| Minimum Detection Limit: | 62.5 pg/mL |
| Application: | ELISA |

Product Details

| | |
|------------------------|--|
| Analytical Method: | Quantitative |
| Detection Method: | Colorimetric |
| Sensitivity: | < 2 pg/mL |
| Components: | 1. One 96-well plate pre-coated with anti-Human P-Cadherin antibody 2. Lyophilized human P-Cadherin standards: 2 tubes (10ng / tube) 3. Sample / Standard diluent buffer: 30ml 4. Biotin conjugated anti-human P-Cadherin antibody (Concentrated): 130 µl. |
| Material not included: | 1. 37 °C incubator 2. Microplate reader (wavelength: 450nm) 3. Precise pipette and disposable pipette tips 4. Automated plate washer 5. ELISA shaker 6. 1.5ml of Eppendorf tubes 7. Plate cover 8. Absorbent filter papers 9. Plastic or glass container with volume of above 1L |

Target Details

| | |
|---------|-------------------|
| Target: | P-Cadherin (CDH3) |
|---------|-------------------|

Target Details

Alternative Name: P-Cadherin ([CDH3 Products](#))

Background: Cadherin-3, also known as P-Cadherin, is a protein that in humans is encoded by the CDH3 gene. This gene is a classical cadherin from the cadherin superfamily. It is located in a six-cadherin cluster in a region on the long arm of chromosome 16 that is involved in loss of heterozygosity events in breast and prostate cancer. This protein is a calcium-dependent cell-cell adhesion glycoprotein composed of five extracellular cadherin repeats, a transmembrane region and a highly conserved cytoplasmic tail. In addition, aberrant expression of this protein is observed in cervical adenocarcinomas. Mutations in this gene have been associated with congenital hypotrichosis with juvenile macular dystrophy.

Application Details

Comment: This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti-P-Cadherin polyclonal antibody was pre-coated onto 96-well plates. And the biotin conjugated anti-P-Cadherin polyclonal antibody was used as detection antibodies. The standards test samples and biotin conjugated detection antibody were added - the wells subsequently and wash with wash buffer. Avidin-Biotin-Peroxidase Complex was added and unbound conjugates were washed away with wash buffer. TMB substrates were used - visualize HRP enzymatic reaction. TMB was catalyzed by HRP - produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional - the P-Cadherin amount of sample captured in plate. Read the O.D. absorbance at 450 nm in a microplate reader and then the concentration of P-Cadherin can be calculated.

Plate: Pre-coated

Reagent Preparation: 1. Before the experiment, centrifuge each kit component for several minutes to bring down all reagents to the bottom of tubes. 2. It is recommend to measure each standard and sample in duplicate. 3. Do NOT let the plate completely dry at any time! Since the dry condition can inactivate the biological material on the plate. 4. Do not reuse pipette tips and tubes to avoid cross contamination. 5. Do not use the expired components and the components from different batches. 6. To avoid the marginal effect of plate incubation for temperature differences (the marginal wells always get stronger reaction), it is recommend to equilibrate the ABC working solution and TMB substrate for at least 30 min at room temperature (37°C) before adding to wells. The TMB substrate (Kit Component 8) is colorless and transparent before use, if not, please contact us for replacement.

Sample Preparation: Preparation of sample and reagents 1. Sample Isolate the test samples soon after collecting, then, analyze immediately (within 2 hours). Or aliquot and store at -20 °C for long term. Avoid

Application Details

multiple freeze-thaw cycles.

Tissue lysate or body fluids, cell culture supernate: Centrifuge to remove precipitate, analyze immediately or aliquot and store at -20 °C .

Serum: Coagulate the serum at room temperature (about 4 hours). Centrifuge at approximately 1000 × g for 10 min. Analyze the serum immediately or aliquot and store at -20 °C .

Plasma: Collect plasma with heparin, citrate or EDTA as the anticoagulant. Centrifuge for 10 min at 1000 x g within 30 min of collection. Analyze immediately or aliquot and store frozen at -20 °C. Note: 1. Coagulate blood samples completely, then, centrifuge, and avoid hemolysis and particle. 2. NaN₃ can not be used as test sample preservative, since it is the inhibitor for HRP.

Restrictions: For Research Use only

Handling

Preservative: Sodium azide, Thimerosal (Merthiolate)