

Datasheet for ABIN1440232

C9 ELISA Kit[Go to Product page](#)**1** Image

Overview

Quantity: 96 tests

Target: C9

Reactivity: Human

Method Type: Sandwich ELISA

Minimum Detection Limit: ~ 0.3 ng/mL

Application: ELISA

Product Details

Purpose: The AssayMax Human C9 ELISA kit is designed for detection of C9 in human plasma, serum, saliva, milk, urine, and cell culture supernatants. This assay employs a quantitative sandwich enzyme immunoassay technique that measures C9 in less than 4 hours. A polyclonal antibody specific for C9 has been pre-coated onto a microplate. C9 in standards and samples is sandwiched by the immobilized antibody and a biotinylated polyclonal antibody specific for C9, which is recognized by a streptavidin-peroxidase conjugate. All unbound material is then washed away and a peroxidase enzyme substrate is added. The color development is stopped and the intensity of the color is measured.

Brand: AssayMax**Sample Type:** Serum, Milk, Saliva, Urine, Plasma, Cell Culture Supernatant**Analytical Method:** Quantitative**Detection Method:** Colorimetric**Cross-Reactivity (Details):** Cross-Reactivity: Monkey 50%, Mouse 1%, Swine 1% Canine 0.5%

Product Details

Characteristics:	Standard Added Value: 0.6 - 6 ng/mL
Components:	<p>Human C9 Microplate: A 96-well polystyrene microplate (12 strips of 8 wells) coated with a polyclonal antibody against human C9.</p> <p>Sealing Tapes: Each kit contains 3 pre-cut, pressure-sensitive sealing tapes that can be cut to fit the format of the individual assay.</p> <p>Human C9 Standard: Human C9 in a buffered protein base (30 ng, lyophilized). AssayMax Human Complement C9 ELISA Kit Catalog No. EC9101-1 This protocol serves as an example for the above. Do not use this protocol in conjunction with any purchased kit. 2</p> <p>Biotinylated C9 Antibody (50x): A 50-fold biotinylated polyclonal antibody against human C9 (140 µL).</p> <p>EIA Diluent Concentrate (10x): A 10-fold concentrated buffered protein base (30 mL).</p> <p>Wash Buffer Concentrate (20x): A 20-fold concentrated buffered surfactant (30 mL, 2 bottles).</p> <p>Streptavidin-Peroxidase Conjugate (SP Conjugate): A 100-fold concentrate (80 µL).</p> <p>Chromogen Substrate: A ready-to-use stabilized peroxidase chromogen substrate tetramethylbenzidine (8 mL).</p> <p>Stop Solution: A 0.5 N hydrochloric acid to stop the chromogen substrate reaction (12 mL).</p>
Material not included:	<p>Microplate reader capable of measuring absorbance at 450 nm.</p> <p>Pipettes (1-20 µL, 20-200 µL, 200-1000µL and multiple channel).</p> <p>Deionized or distilled reagent grade water.</p>

Target Details

Target:	C9
Alternative Name:	Complement C9 (C9 Products)
Background:	<p>Human Complement component 9 (C9) is the terminal component of the complement cascade. It is secreted as an amphiphilic single-chain glycoprotein with 537 amino acids and 71 kDa, and circulates in the blood. The protease alpha-thrombin cleaves C9 at 294 amino acid residues from the carboxy-terminal end and produces two single-chain polypeptides: a hydrophilic C9a and a hydrophobic C9b. In the presence of membrane bound components C5b-8, C9 inserts into the phospholipid bilayer and becomes a pore-forming subunit of the membrane attack complex (MAC) on target membranes. C9-deficient individuals have a significantly increased risk of developing meningococcal meningitis.</p>
Pathways:	Complement System

Application Details

Application Notes: Suggested dilution 1:8000 for Plasma/Serum

Assay Time: < 4 h

Plate: Pre-coated

Reagent Preparation: Freshly dilute all reagents and bring all reagents to room temperature before use.

EIA Diluent Concentrate (10x): If crystals have formed in the concentrate, mix gently until the crystals have completely dissolved. Dilute the EIA Diluent 1:10 with reagent grade water. Store for up to 1 month at 2-8°C.

Standard Curve: Reconstitute the 30 ng of C9 Standard with 2 mL of EIA Diluent to generate a solution of 15 ng/mL. Allow the standard to sit for 10 minutes with gentle agitation prior to making dilutions. Prepare duplicate or triplicate standard points by serially diluting the standard solution (15 ng/mL) 1:2 with EIA Diluent to produce 7.5, 3.75, 1.87, 0.937, 0.468, and 0.234 ng/mL solutions. EIA Diluent serves as the zero standard (0 ng/mL). Any remaining solution should be frozen at -20°C and used within 30 days.

Biotinylated C9 Antibody (50x): Spin down the antibody briefly and dilute the desired amount of the antibody 1:50 with EIA Diluent. Any remaining solution should be frozen at -20°C.

Wash Buffer Concentrate (20x): If crystals have formed in the concentrate, mix gently until the crystals have completely dissolved. Dilute the Wash Buffer Concentrate 1:20 with reagent grade water.

SP Conjugate (100x): Spin down the SP Conjugate briefly and dilute the desired amount of the conjugate 1:100 with EIA Diluent. Any remaining solution should be frozen at -20°C.

Sample Preparation: Plasma: Collect plasma using one-tenth volume of 0.1 M sodium citrate as an anticoagulant. Centrifuge samples at 2000 x g for 10 minutes and assay. Dilute samples 1:20000 into EIA Diluent Store samples at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles. (EDTA or Heparin can also be used as an anticoagulant.)

Serum: Samples should be collected into a serum separator tube. After clot formation, centrifuge samples at 2000 x g for 10 minutes. Dilute samples 1:20000 into EIA Diluent. Store samples at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles.

Cell Culture Supernatants: Centrifuge cell culture media at 2000 x g for 10 minutes to remove debris. Collect supernatants and assay. Store the remaining samples at -20°C or below. Avoid repeated freeze-thaw cycles.

Milk: Collect milk using sample tube. Centrifuge samples at 600 x g for 10 minutes and assay. Dilute samples 1:50 into EIA Diluent. Store samples at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles.

Saliva: Collect saliva using sample tube. Centrifuge samples at 600 x g for 10 minutes and

assay. Store undiluted samples at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles.

Urine: Collect urine using sample pot. Centrifuge samples at 800 x g for 10 minutes and assay. Store undiluted samples at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles.

Assay Procedure:

Prepare all reagents, working standards and samples as instructed. Bring all reagents to room temperature before use. The assay is performed at room temperature (20-30°C).

Remove excess microplate strips from the plate frame and return them immediately to the foil pouch with desiccant inside. Reseal the pouch securely to minimize exposure to water vapor and store in a vacuum desiccator.

Add 50 µL of Standard or sample per well. Cover wells with a sealing tape and incubate for two hours. Start the timer after the last sample addition.

Wash five times with 200 µL of Wash Buffer manually. Invert the plate each time and decant the contents, hit it 4-5 times on absorbent paper towel to completely remove the liquid. If using a machine wash six times with 300 µL of Wash Buffer and then invert the plate, decant the contents, hit it 4-5 times on absorbent paper towel to completely remove the liquid.

Add 50 µL of Biotinylated C9 Antibody to each well and incubate for one hour.

Wash the microplate as described above.

Add 50 µL of Streptavidin-Peroxidase Conjugate per well and incubate for 30 minutes. Turn on the microplate reader and set up the program in advance.

Wash the microplate as described above.

Add 50 µL of Chromogen Substrate per well and incubate for about 10 minutes or till the optimal blue color density develops. Gently tap the plate to ensure thorough mixing and break the bubbles in the well with pipette tip.

Add 50 µL of Stop Solution to each well. The color will change from blue to yellow.

Read the absorbance on a microplate reader at a wavelength of 450 nm immediately. If wavelength correction is available, subtract readings at 570 nm from those at 450 nm to correct optical imperfections. Otherwise, read the plate at 450 nm only. Please note that some unstable black particles may be generated at high concentration points after stopping the reaction for about 10 minutes, which will reduce the readings.

Calculation of Results:

Calculate the mean value of the duplicate or triplicate readings for each standard and sample.

To generate a standard curve, plot the graph using the standard concentrations on the x-axis and the corresponding mean 450 nm absorbance on the y-axis. The best-fit line can be determined by regression analysis using four-parameter or log-log logistic curve-fit.

Determine the unknown sample concentration from the Standard Curve and multiply the value

Application Details

by the dilution factor.

Assay Precision: Intra-assay and inter-assay coefficients of variation were 4.3% and 7.0% respectively.

Restrictions: For Research Use only

Handling

Handling Advice: Prepare all reagents (working diluent buffer, wash buffer, standards, biotinylated- antibody, and SP conjugate) as instructed, prior to running the assay.

Prepare all samples prior to running the assay. The dilution factors for the samples are suggested in this protocol. However, the user should determine the optimal dilution factor.

Spin down the SP conjugate vial and the biotinylated-antibody vial before opening and using contents.

The kit should not be used beyond the expiration date.

The Stop Solution is an acid solution.

Storage: 4 °C/-20 °C

Storage Comment: Store components of the kit at 2-8°C or -20°C upon arrival up to the expiration date.

Store SP Conjugate and Biotinylated Antibody at -20°C

Store Microplate, Diluent Concentrate (10x), Wash Buffer, Stop Solution, and Chromogen Substrate at 2-8°C


Opened unused microplate wells may be returned to the foil pouch with the desiccant packs.

Reseal along zip-seal. May be stored for up to 1 month in a vacuum desiccator.

Diluent (1x) may be stored for up to 1 month at 2-8°C.

Store Standard at 2-8°C before reconstituting with Diluent and at -20°C after reconstituting with Diluent.

**Human Complement C9
Standard Curve**



ELISA

Image 1.