antibodies -online.com







C5A ELISA Kit





Overview

Quantity:	96 tests
Target:	C5A (C5a)
Binding Specificity:	AA 679-755
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Detection Range:	15.6-1000 pg/mL
Minimum Detection Limit:	15.6 pg/mL
Application:	ELISA

Product Details

Purpose:	Sandwich High Sensitivity ELISA kit for Quantitative Detection of Mouse Complement C5a
Brand:	PicoKine™
Sample Type:	Cell Culture Supernatant, Serum, Plasma (heparin), Plasma (EDTA)
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Immunogen:	Expression system for standard: E.coli
	Immunogen sequence: N679-R755
Specificity:	Expression system for standard: E.coli
	Immunogen sequence: N679-R755
Cross-Reactivity (Details):	There is no detectable cross-reactivity with other relevant proteins.

Product Details

Sensitivity:	<10pg/mL
Material not included:	Microplate reader in standard size. Automated plate washer. Adjustable pipettes and pipette tips. Multichannel pipettes are recommended in the condition of large amount of samples in the detection. Clean tubes and Eppendorf tubes. Washing buffer (neutral PBS or TBS). Preparation of 0.01M TBS: Add 1.2g Tris, 8.5g Nacl
Target Details	
Target:	C5A (C5a)
Alternative Name:	Complement C5a (C5a Products)
Background:	Protein Function: Activation of C5 by a C5 convertase initiates the spontaneous assembly of the late complement components, C5-C9, into the membrane attack complex. C5b has a transient binding site for C6. The C5b-C6 complex is the foundation upon which the lytic complex is assembled. Background: C5a is a protein fragment released from complement component C5. This gene is mapped to 9q33.2. The protein encoded by this gene is the fifth component of complement, which plays an important role in inflammatory and cell killing processes. This protein is comprised of alpha and beta polypeptide chains that are linked by a disulfide bridge. An activation peptide, C5a, which is an anaphylatoxin that possesses potent spasmogenic and chemotactic activity, is derived from the alpha polypeptide via cleavage with a convertase. The C5b macromolecular cleavage product can form a complex with the C6 complement component, and this complex is the basis for formation of the membrane attack complex, which includes additional complement components. Mutations in this gene cause complement component 5 deficiency, a disease where patients show a propensity for severe recurrent infections. Defects in this gene have also been linked to susceptibility to liver fibrosis and to rheumatoid arthritis. Synonyms: Complement C5, Hemolytic complement, Complement C5 beta chain, Complement C5 alpha chain, C5a anaphylatoxin, Complement C5 alpha' chain, C5, Hc, Full Gene Name: Complement C5 Cellular Localisation: Secreted.
Gene ID:	15139
UniProt:	P06684
Pathways:	Complement System, Carbohydrate Homeostasis

Application Details


Application Notes:	Before using Kit, spin tubes and bring down all components to bottom of tube. Duplicate well assay was recommended for both standard and sample testing.
Comment:	Sequence similarities: Contains 1 anaphylatoxin-like domain.
Plate:	Pre-coated
Protocol:	mouse Complement C5a ELISA Kit was based on standard sandwich enzyme-linked immune-sorbent assay technology. A monoclonal antibody from rat specific for Complement C5a has been precoated onto 96-well plates. Standards(E.coli, N679-R755) and test samples are added to the wells, a biotinylated detection polyclonal antibody from goat specific for Complement C5a is added subsequently and then followed by washing with PBS or TBS buffer. Avidin-Biotin-Peroxidase Complex was added and unbound conjugates were washed away with PBS or TBS buffer. HRP substrate TMB was used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the mouse Complement C5a amount of sample captured in plate.
Assay Procedure:	Aliquot 0.1 mL per well of the 1000pg/mL, 500pg/mL, 250pg/mL, 125pg/mL, 62.5pg/mL, 31.2pg/mL, 15.6pg/mL mouse Complement C5a standard solutions into the precoated 96-well plate. Add 0.1 mL of the sample diluent buffer into the control well (Zero well). Add 0.1 mL of each properly diluted sample of mouse cell culture supernates, serum or plasma(heparin, EDTA) to each empty well. See "Sample Dilution Guideline" above for details. It is recommended that each mouse Complement C5a standard solution and each sample be measured in duplicate.
Assay Precision:	 Sample 1: n=16, Mean(pg/ml): 37.7, Standard deviation: 1.39, CV(%): 3.7 Sample 2: n=16, Mean(pg/ml): 118, Standard deviation: 5.43, CV(%): 4.6 Sample 3: n=16, Mean(pg/ml): 621, Standard deviation: 32.91, CV(%): 5.3, Sample 1: n=24, Mean(pg/ml): , Standard deviation: , CV(%): Sample 2: n=24, Mean(pg/ml): , Standard deviation: , CV(%): Sample 3: n=24, Mean(pg/ml): , Standard deviation: , CV(%):
Restrictions:	For Research Use only
Handling	
Handling Advice:	Avoid multiple freeze-thaw cycles.
Storage:	-20 °C,4 °C

Expiry Date:

12 months

Images

Mouse Complement C5a ELISA Kit



ELISA

Image 1. Mouse Complement C5a PicoKine ELISA Kit standard curve