# antibodies .- online.com







### **FGF1 ELISA Kit**





#### Overview

Quantity:	96 tests
Target:	FGF1
Binding Specificity:	AA 16-155
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Detection Range:	31.2-2000 pg/mL
Minimum Detection Limit:	31.2 pg/mL
Application:	ELISA

#### **Product Details**


Purpose:	Sandwich High Sensitivity ELISA kit for Quantitative Detection of Mouse FGF1
Brand:	PicoKine™
Sample Type:	Cell Culture Supernatant, Serum, Plasma (heparin), Plasma (EDTA)
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Immunogen:	Expression system for standard: E.coli Immunogen sequence: F16-D155
Specificity:	Expression system for standard: E.coli Immunogen sequence: F16-D155
Cross-Reactivity (Details):	There is no detectable cross-reactivity with other relevant proteins.

#### **Product Details**

Sensitivity:	<10pg/mL
Material not included:	Microplate reader in standard size. Automated plate washer. Adjustable pipettes and pipette
	tips. Multichannel pipettes are recommended in the condition of large amount of samples in the
	detection. Clean tubes and Eppendorf tubes. Washing buffer (neutral PBS or TBS). Preparation
	of 0.01M TBS: Add 1.2g Tris, 8.5g Nacl
Target Details	
Target:	FGF1
Alternative Name:	FGF1 (FGF1 Products)
Background:	Protein Function: Plays an important role in the regulation of cell survival, cell division,
	angiogenesis, cell differentiation and cell migration. Functions as potent mitogen in vitro (By similarity)
	Background: Heparin-binding growth factor 1 is a protein that in humans is encoded by the
	FGF1 gene. The protein encoded by this gene is a member of the fibroblast growth factor(FGF)
	family. This protein functions as a modifier of endothelial cell migration and proliferation, as
	well as an angiogenic factor. It acts as a mitogen for a variety of mesoderm- and
	neuroectoderm-derived cells in vitro, thus is thought to be involved in organogenesis. The FGF1
	gene was mapped to chromosome 5q31.3-q33.2 by in situ hybridization.
	Synonyms: Fibroblast growth factor 1,FGF-1,Acidic fibroblast growth factor,aFGF,Heparin-
	binding growth factor 1,HBGF-1,Fgf1,Fgf-1, Fgfa,
	Full Gene Name: Fibroblast growth factor 1
	Cellular Localisation: Secreted. Cytoplasm . Cytoplasm, cell cortex . Cytoplasm, cytosol .
	Nucleus . Lacks a cleavable signal sequence. Within the cytoplasm, it is transported to the cell
	membrane and then secreted by a non-classical pathway that requires Cu(2+) ions and
	S100A13. Secreted in a complex with SYT1. Binding of exogenous FGF1 to FGFR facilitates
	endocytosis followed by translocation of FGF1 across endosomal membrane into the cytosol.
	Nuclear import from the cytosol requires the classical nuclear import machinery, involving
	proteins KPNA1 and KPNB1, as well as LRRC59 (By similarity)
Gene ID:	14164
UniProt:	P61148
Pathways:	RTK Signaling, Fc-epsilon Receptor Signaling Pathway, EGFR Signaling Pathway, Neurotrophin
	Signaling Pathway

## **Application Details**

Application Notes:	Before using Kit, spin tubes and bring down all components to bottom of tube. Duplicate well
	assay was recommended for both standard and sample testing.
Plate:	Pre-coated
Protocol:	mouse FGF1 ELISA Kit was based on standard sandwich enzyme-linked immune-sorbent assa
	technology. A monoclonal antibody from rat specific for FGF1 has been precoated onto 96-we
	plates. Standards(E.coli, F16-D155) and test samples are added to the wells, a biotinylated
	detection polyclonal antibody from goat specific for FGF1 is added subsequently and then
	followed by washing with PBS or TBS buffer. Avidin-Biotin-Peroxidase Complex was added and
	unbound conjugates were washed away with PBS or TBS buffer. HRP substrate TMB was used
	to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color
	product that changed into yellow after adding acidic stop solution. The density of yellow is
	proportional to the mouse FGF1 amount of sample captured in plate.
Assay Procedure:	Aliquot 0.1 mL per well of the 2000pg/mL,1000pg/mL, 500pg/mL, 250pg/mL, 125pg/mL,
,	62.5pg/mL, 31.2pg/mL mouse FGF1 standard solutions into the precoated 96-well plate. Add
	0.1 mL of the sample diluent buffer into the control well (Zero well). Add 0.1 mL of each
	properly diluted sample of mouse cell culture supernates, serum or plasma(heparin, EDTA) to
	each empty well. See "Sample Dilution Guideline" above for details. It is recommended that
	each mouse FGF1 standard solution and each sample be measured in duplicate.
Assay Precision:	• Sample 1: n=16, Mean(pg/ml): 158, Standard deviation: 8.53, CV(%): 5.4
	<ul> <li>Sample 2: n=16, Mean(pg/ml): 754, Standard deviation: 58.81, CV(%): 7.8</li> <li>Sample 3: n=16, Mean(pg/ml): 1219, Standard deviation: 41.45, CV(%): 3.4,</li> </ul>
	<ul> <li>Sample 3. 11–10, Mean(pg/ml): 1219, Standard deviation: 41.43, CV(%): 3.4,</li> <li>Sample 1: n=24, Mean(pg/ml): 152, Standard deviation: 13.98, CV(%): 9.2</li> </ul>
	• Sample 2: n=24, Mean(pg/ml): 747, Standard deviation: 63.50, CV(%): 8.5
	Sample 3: n=24, Mean(pg/ml): 1223, Standard deviation: 111.3, CV(%): 9.1
Restrictions:	For Research Use only
Handling	
Handling Advice:	Avoid multiple freeze-thaw cycles.
Storage:	-20 °C,4 °C
Storage Comment:	Store at 4°C for 6 months, at -20°C for 12 months. Avoid multiple freeze-thaw cycles
Expiry Date:	12 months



#### **ELISA**

Image 1. Mouse FGF1 PicoKine ELISA Kit standard curve