antibodies - online.com







TWEAK ELISA Kit





Overview

Quantity:	96 tests
Target:	TWEAK (TNFSF12)
Binding Specificity:	AA 94-249
Reactivity:	Rat
Method Type:	Sandwich ELISA
Detection Range:	15.6-1000 pg/mL
Minimum Detection Limit:	15.6 pg/mL
Application:	ELISA

Product Details

Purpose:	Sandwich High Sensitivity ELISA kit for Quantitative Detection of Rat TNFSF12/TWEAK
Brand:	PicoKine™
Sample Type:	Cell Culture Supernatant, Serum, Plasma (heparin), Plasma (EDTA)
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Immunogen:	Expression system for standard: E.coli Immunogen sequence: S94-H249
Specificity:	Expression system for standard: E.coli Immunogen sequence: S94-H249
Cross-Reactivity (Details):	There is no detectable cross-reactivity with other relevant proteins.

Product Details

Sensitivity:	<10pg/mL
Material not included:	Microplate reader in standard size. Automated plate washer. Adjustable pipettes and pipette tips. Multichannel pipettes are recommended in the condition of large amount of samples in the detection. Clean tubes and Eppendorf tubes. Washing buffer (neutral PBS or TBS). Preparation of 0.01M TBS: Add 1.2g Tris, 8.5g Nacl
Target Details	
Target:	TWEAK (TNFSF12)
Alternative Name:	TNFSF12 (TNFSF12 Products)
Background:	Background: Tumor necrosis factor ligand superfamily member 12 also known as TNF-related weak inducer of apoptosis(TWEAK) is a protein that in humans is encoded by the TNFSF12 gene. The protein encoded by this gene is a cytokine that belongs to the tumor necrosis factor(TNF) ligand family. This protein is a ligand for the FN14/TWEAKR receptor. This cytokine has overlapping signaling functions with TNF, but displays a much wider tissue distribution. This cytokine can induce apoptosis via multiple pathways of cell death in a cell type-specific manner. This cytokine is also found to promote proliferation and migration of endothelial cells, and thus acts as a regulator ofangiogenesis. The TNFSF12 gene lies 878 bp upstream of the putative transcriptional start site of the TNFSF13 gene on chromosome 17p13.1. Synonyms: Protein Prmt2 ,Tumor necrosis factor ligand superfamily member 12 ,Tnfsf12 ,Prmt2 , Full Gene Name: tumor necrosis factor (ligand) superfamily, member 12
Gene ID:	360548
UniProt:	Q6AYC1
Pathways:	Apoptosis
Application Details	
Application Notes:	Before using Kit, spin tubes and bring down all components to bottom of tube. Duplicate well assay was recommended for both standard and sample testing.
Plate:	Pre-coated
Protocol:	rat TNFSF12 ELISA Kit was based on standard sandwich enzyme-linked immune-sorbent assay technology. A monoclonal antibody from mouse specific for TNFSF12 has been precoated onto 96-well plates. Standards(E.coli, S94-H249) and test samples are added to the wells, a

biotinylated detection polyclonal antibody from goat specific for TNFSF12 is added
subsequently and then followed by washing with PBS or TBS buffer. Avidin-Biotin-Peroxidase
$ {\it Complex was added and unbound conjugates were washed away with PBS or TBS buffer. \ HRP } \\$
substrate TMB was used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to
produce a blue color product that changed into yellow after adding acidic stop solution. The
density of yellow is proportional to the rat TNFSF12 amount of sample captured in plate.

Assay Procedure:

Aliquot 0.1 mL per well of the 1000pg/mL, 500pg/mL, 250pg/mL, 125pg/mL, 62.5pg/mL, 31.2pg/mL, 15.6pg/mL rat TNFSF12 standard solutions into the precoated 96-well plate. Add 0.1 mL of the sample diluent buffer into the control well (Zero well). Add 0.1 mL of each properly diluted sample of rat cell culture supernates, serum or plasma(heparin, EDTA) to each empty well. See "Sample Dilution Guideline" above for details. It is recommended that each rat TNFSF12 standard solution and each sample be measured in duplicate.

Assay Precision:


- Sample 1: n=16, Mean(pg/ml): 73, Standard deviation: 2.7, CV(%): 3.7
- Sample 2: n=16, Mean(pg/ml): 236, Standard deviation: 10.62, CV(%): 4.5
- Sample 3: n=16, Mean(pg/ml): 533, Standard deviation: 28.8, CV(%): 5.4,
- Sample 1: n=24, Mean(pg/ml): 88, Standard deviation: 5.2, CV(%): 4.8
- Sample 2: n=24, Mean(pg/ml): 348, Standard deviation: 19.5, CV(%): 5.6
- Sample 3: n=24, Mean(pg/ml): 635, Standard deviation: 38.74, CV(%): 6.1

Restrictions:

For Research Use only

Handling

Handling Advice:	Avoid multiple freeze-thaw cycles.
Storage:	-20 °C,4 °C
Storage Comment:	Store at 4°C for 6 months, at -20°C for 12 months. Avoid multiple freeze-thaw cycles
Expiry Date:	12 months



ELISA

Image 1. Rat TNFSF12/TWEAK PicoKine ELISA Kit standard curve