# antibodies - online.com







# **LUM ELISA Kit**





# Overview

Quantity:	96 tests
Target:	LUM
Binding Specificity:	AA 19-338
Reactivity:	Rat
Method Type:	Sandwich ELISA
Detection Range:	156-10.000 pg/mL
Minimum Detection Limit:	156 pg/mL
Application:	ELISA

## **Product Details**

Purpose:	Sandwich High Sensitivity ELISA kit for Quantitative Detection of Rat Lumican
Brand:	PicoKine™
Sample Type:	Cell Culture Supernatant, Serum, Plasma (heparin)
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Immunogen:	Expression system for standard: NSO
	Immunogen sequence: Q19-N338
Specificity:	Expression system for standard: NSO
	Immunogen sequence: Q19-N338
Cross-Reactivity (Details):	There is no detectable cross-reactivity with other relevant proteins.

## **Product Details**

Sensitivity:	<10pg/mL
Material not included:	Microplate reader in standard size. Automated plate washer. Adjustable pipettes and pipette tips. Multichannel pipettes are recommended in the condition of large amount of samples in the detection. Clean tubes and Eppendorf tubes. Washing buffer (neutral PBS or TBS). Preparation of 0.01M TBS: Add 1.2g Tris, 8.5g Nacl
Target Details	
Target:	LUM
Alternative Name:	LUM (LUM Products)
Background:	Background: Lumican, also known as LUM, is a protein which in humans is encoded by the LUM gene. This gene is a member of the small interstitial proteoglycan gene(SIPG) family, and it is a keratan sulfate proteoglycan which presents large quantities in the corneal stroma and in interstitial collagenous matrices of the heart, aorta, skeletal muscle, skin, and intervertebral discs. Lumican is mapped to 12q21.33, it interacts with collagen and limits growth of fibrils in diameter. In the cornea, Lumican not only interacts with collagen molecules to limit fibril growth, but also plays a critical role in the regular spacing of fibrils and acquisition of corneal transparency by virtue of its keratan sulfate-containing glycosaminoglycan side chains LDC. Synonyms: Lumican,Keratan sulfate proteoglycan lumican,KSPG lumican,Lum,Lcn, Ldc, Full Gene Name: Lumican  Cellular Localisation: Secreted, extracellular space, extracellular matrix.
Gene ID:	81682
UniProt:	P51886
Pathways:	Glycosaminoglycan Metabolic Process
Application Details	
Application Notes:	Before using Kit, spin tubes and bring down all components to bottom of tube. Duplicate well assay was recommended for both standard and sample testing.
Comment:	Sequence similarities: Belongs to the small leucine-rich proteoglycan (SLRP) family. SLRP class II subfamily.
Plate:	Pre-coated
Protocol:	rat Lumican ELISA Kit was based on standard sandwich enzyme-linked immune-sorbent assay technology. A monoclonal antibody from mouse specific for Lumican has been precoated onto

96-well plates. Standards(NSO, Q19-N338) and test samples are added to the wells, a
biotinylated detection polyclonal antibody from goat specific for Lumican is added
subsequently and then followed by washing with PBS or TBS buffer. Avidin-Biotin-Peroxidase
Complex was added and unbound conjugates were washed away with PBS or TBS buffer. HRP
substrate TMB was used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to
produce a blue color product that changed into yellow after adding acidic stop solution. The
density of yellow is proportional to the rat Lumican amount of sample captured in plate.

#### Assay Procedure:

Aliquot 0.1 mL per well of the 10,000pg/mL, 5000pg/mL, 2500pg/mL, 1250pg/mL, 625pg/mL, 312pg/mL, 156pg/mL rat Lumican standard solutions into the precoated 96-well plate. Add 0.1 mL of the sample diluent buffer into the control well (Zero well). Add 0.1 mL of each properly diluted sample of rat cell culture supernates, serum or plasma(heparin) to each empty well. See "Sample Dilution Guideline" above for details. It is recommended that each rat Lumican standard solution and each sample be measured in duplicate.

#### Assay Precision:


- Sample 1: n=16, Mean(pg/ml): 1258, Standard deviation: 59.13, CV(%): 4.7
- Sample 2: n=16, Mean(pg/ml): 3733, Standard deviation: 201.6, CV(%): 5.4
- Sample 3: n=16, Mean(pg/ml): 6233, Standard deviation: 374, CV(%): 6,
- Sample 1: n=24, Mean(pg/ml): 1476, Standard deviation: 87.1, CV(%): 5.9
- Sample 2: n=24, Mean(pg/ml): 3946, Standard deviation: 248.6, CV(%): 6.3
- Sample 3: n=24, Mean(pg/ml): 6852, Standard deviation: 500.2, CV(%): 7.3

Restrictions:

For Research Use only

#### Handling

Handling Advice:	Avoid multiple freeze-thaw cycles.
Storage:	-20 °C,4 °C
Storage Comment:	Store at 4°C for 6 months, at -20°C for 12 months. Avoid multiple freeze-thaw cycles
Expiry Date:	12 months



## **ELISA**

Image 1. Rat Lumican PicoKine ELISA Kit standard curve