

Datasheet for ABIN1979430
BCL2L2 ELISA Kit[Go to Product page](#)

1 Image

Overview

Quantity:	96 tests
Target:	BCL2L2
Reactivity:	Rat
Method Type:	Sandwich ELISA
Detection Range:	0.4-100 ng/mL
Minimum Detection Limit:	0.4 ng/mL
Application:	ELISA

Product Details

Purpose:	Rat bcl-w ELISA Kit for cell culture supernatants, plasma, and serum samples.
Sample Type:	Serum, Plasma, Cell Culture Supernatant
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA antibody pair also can recognize human and mouse BCL-W.
Characteristics:	<ul style="list-style-type: none">• Strip plates and additional reagents allow for use in multiple experiments• Quantitative protein detection• Establishes normal range• The best products for confirmation of antibody array data
Components:	<ul style="list-style-type: none">• Pre-Coated 96-well Strip Microplate• Wash Buffer• Stop Solution

Product Details

- Assay Diluent(s)
- Lyophilized Standard
- Biotinylated Detection Antibody
- Streptavidin-Conjugated HRP
- TMB One-Step Substrate

Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2 µL to 1 µL volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 µL and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target: BCL2L2

Alternative Name: BCL-W ([BCL2L2 Products](#))

Gene ID: 60434

UniProt: [O88996](#)

Application Details

Application Notes: Recommended Dilution for serum and plasma samples 2 fold

Sample Volume: 100 µL

Plate: Pre-coated

Protocol:

1. Prepare all reagents, samples and standards as instructed in the manual.
2. Add 100 µL of standard or sample to each well.
3. Incubate 2.5 h at RT or O/N at 4 °C.
4. Add 100 µL of prepared biotin antibody to each well.
5. Incubate 1 h at RT.
6. Add 100 µL of prepared Streptavidin solution to each well.
7. Incubate 45 min at RT.
8. Add 100 µL of TMB One-Step Substrate Reagent to each well.
9. Incubate 30 min at RT.
10. Add 50 µL of Stop Solution to each well.
11. Read at 450 nm immediately.

Application Details

Reagent Preparation:

1. Bring all reagents and samples to room temperature (18 - 25 °C) before use. 2. Sample dilution: If your samples need to be diluted, Assay Diluent A (Item D) should be used for dilution of serum/plasma samples. 1x Assay Diluent B (Item E) should be used for dilution of culture supernatants and urine. The Rat BCL-W ELISA Kit Protocol 3 Suggested dilution for normal serum/plasma: 2 fold*. * Please note that levels of the target protein may vary between different specimens. Optimal dilution factors for each sample must be determined by the investigator. 3. Assay Diluent B should be diluted 5-fold with deionized or distilled water before use. 4. Preparation of standard: Briefly spin the vial of Item C and then add 800 µL Assay Diluent A (for serum/plasma samples) or 1x Assay Diluent B (for cell culture supernates) into Item C vial to prepare a 100 ng/ml standard. Dissolve the powder thoroughly by a gentle mix. Pipette 300 µL Assay Diluent A or 1x Assay Diluent B into each tube. Use the 100 ng/ml standard solution to produce a dilution series (shown below). Mix each tube thoroughly before the next transfer. Assay Diluent A or 1x Assay Diluent B serves as the zero standard (0 ng/ml).

200 µL Standard, Item C + 800 µL	200µl	200 µL	200 µL	200 µL	200 µL	100	40	16	6.4	2.56	1.024	0.410	0
ng/ml	ng/ml	ng/ml	ng/ml	ng/ml	ng/ml	ng/ml	ng/ml	ng/ml	ng/ml	ng/ml	ng/ml	ng/ml	ng/ml

5. If the Wash Concentrate (20x) (Item B) contains visible crystals, warm to room temperature and mix gently until dissolved. Dilute 20 mL of Wash Buffer Concentrate into deionized or distilled water to yield 400 mL of 1x Wash Buffer. 6. Briefly spin the Detection Antibody vial (Item F) before use. Add 100 µL of 1x Assay Diluent B into the vial to prepare a detection antibody The Rat BCL-W ELISA Kit Protocol 4 concentrate. Pipette up and down to mix gently (the concentrate can be stored at 4 °C for 5 days). The detection antibody concentrate should be diluted 80-fold with 1x Assay Diluent B and used in step 4 of Part VI Assay Procedure. 7. Briefly spin the HRP-Streptavidin concentrate vial (Item G) and pipette up and down to mix gently before use. HRP-Streptavidin concentrate should be diluted 500-fold with 1x Assay Diluent B. For example: Briefly spin the vial (Item G) and pipette up and down to mix gently . Add 20 µL of HRP-Streptavidin concentrate into a tube with 10 mL 1x Assay Diluent B to prepare a 500-fold diluted HRP- Streptavidin solution (don't store the diluted solution for next day use). Mix well.

Assay Procedure:

1. Bring all reagents and samples to room temperature (18 - 25 °C) before use. It is recommended that all standards and samples be run at least in duplicate. 2. Add 100 µL of each standard (see Reagent Preparation step 2) and sample into appropriate wells. Cover well and incubate for 2.5 hours at room temperature or overnight at 4 °C with gentle shaking. 3. Discard 4. Add 100 µL of 1x prepared biotinylated antibody (Reagent Preparation step 6) to each well. Incubate for 1 hour at room temperature with gentle shaking. 5. Discard the solution. Repeat the wash as in step 3. 6. Add 100 µL of prepared Streptavidin solution (see Reagent Preparation step 7) to each well. Incubate for 45 minutes at room temperature with gentle

Application Details

shaking. 7. Discard the solution. Repeat the wash as in step 3. 8. Add 100 µL of TMB One-Step Substrate Reagent (Item H) to each well. Incubate for 30 minutes at room temperature in the dark with gentle shaking. 9. Add 50 µL of Stop Solution (Item I) to each well. Read at 450 nm immediately.

Calculation of Results: Calculate the mean absorbance for each set of duplicate standards, controls and samples, and subtract the average zero standard optical density. Plot the standard curve on log-log graph paper or using Sigma plot software, with standard concentration on the x-axis and absorbance on the y-axis. Draw the best-fit straight line through the standard points.

Assay Precision: Intra-Assay: CV<10%
Inter-Assay: CV<12%

Restrictions: For Research Use only

Handling

Handling Advice: Avoid repeated freeze-thaw cycles.

Storage: -20 °C

Storage Comment: The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C.

Expiry Date: 6 months

Images

ELISA

Image 1.

