

Datasheet for ABIN2039193
C19ORF80 ELISA Kit



[Go to Product page](#)

1 Image **8** Publications

Overview

Quantity:	96 tests
Target:	C19ORF80
Reactivity:	Human
Detection Range:	6.25-400 pg/mL
Minimum Detection Limit:	6.25 pg/mL
Application:	ELISA

Product Details

Sample Type:	Serum, Plasma, Tissue Homogenate
Detection Method:	Colorimetric
Cross-Reactivity (Details):	Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between the target antigen and all analogues for other species. Therefore, cross reaction may still exist.
Sensitivity:	1.56 pg/mL
Components:	<ul style="list-style-type: none"> • Assay plate (12 × 8 coated Microwells) • Standard (freeze dried) • Biotin-antibody (100 × concentrate) • HRP-avidin (100 × concentrate) • Biotin-antibody Diluent • HRP-avidin Diluent • Sample Diluent • Wash Buffer (25 × concentrate) • TMB Substrate

Product Details

- Stop Solution
- Adhesive Strip (for 96 wells)
- Instruction manual

Target Details

Target:	C19ORF80
Alternative Name:	Angiopoietin-Like Protein 8 (Angptl8) (C19ORF80 Products)
Background:	Synonyms: Betatrophin,Lipasin,C19orf80,UNQ599,PRO1185
UniProt:	Q6UXH0

Application Details

Application Notes:	<ul style="list-style-type: none">• The supplier is only responsible for the kit itself, but not for the samples consumed during the assay. The user should calculate the possible amount of the samples used in the whole test. Please reserve sufficient samples in advance.• Samples to be used within 5 days may be stored at 2-8°C, otherwise samples must be stored at -20°C (\leq 1 month) or -80°C (\leq 2 months) to avoid loss of bioactivity and contamination.• Grossly hemolyzed samples are not suitable for use in this assay.• If the samples are not indicated in the manual, a preliminary experiment to determine the validity of the kit is necessary.• Please predict the concentration before assaying. If values for these are not within the range of the standard curve, users must determine the optimal sample dilutions for their particular experiments.• Tissue or cell extraction samples prepared by chemical lysis buffer may cause unexpected ELISA results due to the impacts of certain chemicals.• Owing to the possibility of mismatching between antigens from another resource and antibodies used in this supplier's kits (e.g., antibody targets conformational epitope rather than linear epitope), some native or recombinant proteins from other manufacturers may not be recognized by this supplier's products.• Influenced by factors including cell viability, cell number and cell sampling time, samples from cell culture supernatant may not be recognized by the kit.• Fresh samples without long time storage are recommended for the test. Otherwise, protein degradation and denaturalization may occur in those samples and finally lead to wrong results.
--------------------	--

Comment:	Detection wavelength: 450 nm
	Information on standard material: Depending on the antigen to be detected, standards can be either native or recombinant

protein. The recombinant proteins are being expressed in CHO cells in most cases. Please inquire for more information. The formulation of auxiliary material in the standard is considered proprietary information, however it does not contain any poisonous substance. Proclin 300 (1:3000) is used as preservative.

Information on reagents:

In most cases the stop solution provided is 1 N H₂SO₄. The formulation of wash solution is proprietary information. None of the components contain (sodium) azide, thimerosal, 2-mercaptoethanol (2-ME) or any other poisonous materials. For the sandwich method kits, the sample diluent, antibody diluent, enzyme diluent and standard all contain BSA.

Information on antibodies:

The antibodies provided in different kits vary in regards to clonality and host. Some antibodies are affinity purified, some are Protein A

Assay Time:	1 - 4.5 h
Plate:	Pre-coated
Restrictions:	For Research Use only

Handling

Precaution of Use:	The Stop Solution provided with this kit is an acid solution. Wear eye, hand, face and clothing protection when using this material.
Handling Advice:	<ul style="list-style-type: none">• The kit should not be used beyond the expiration date on the kit label.• Do not mix or substitute reagents with those from other lots or sources.• If samples generate values higher than the highest standard, dilute the samples with Sample Diluent and repeat the assay.• Any variation in Sample Diluent, operator, pipetting technique, washing technique, incubation time/temperature and kit age can cause variation in binding.• This assay is designed to eliminate interference by soluble receptors, binding proteins and other factors present in biological samples. Until all factors have been tested in the Immunoassay, the possibility of interference cannot be excluded.
Storage:	4 °C/-20 °C
Storage Comment:	For unopened kit: All the reagents should be kept according to the labels on vials.
Expiry Date:	6 months

Publications

Product cited in: Pascual-Corrales, Gómez-Ambrosi, Moncada, Valentí, Catalán, Rodríguez, Ramírez, Silva, Gil, Salvador, Frühbeck: "Circulating ANGPTL8/Betatrophin Concentrations Are Increased After Surgically Induced Weight Loss, but Not After Diet-Induced Weight Loss." in: **Obesity surgery**, Vol. 26, Issue 8, pp. 1881-9, (2018) ([PubMed](#)).

Abu-Farha, Abubaker, Tuomilehto: "ANGPTL8 (betatrophin) role in diabetes and metabolic diseases." in: **Diabetes/metabolism research and reviews**, Vol. 33, Issue 8, (2018) ([PubMed](#)).


Calan, Yilmaz, Kume, Unal Kocabas, Yesil Senses, Senses, Temur, Gursoy Calan: "Elevated circulating levels of betatrophin are associated with polycystic ovary syndrome." in: **Endocrine**, Vol. 53, Issue 1, pp. 271-9, (2017) ([PubMed](#)).

Ejarque, Borlaug, Vilarrasa, Martinez-Perez, Llauradó, Megía, Helland, Gutierrez, Serena, Folkestad, Nuñez-Roa, Roche, Casajoana, Fradera, González-Clemente, López, Mohn, Nedrebø, Nogueiras, Mellgren et al.: "Angiopoietin-like protein 8/betatrophin as a new determinant of type 2 diabetes remission after bariatric surgery. ..." in: **Translational research : the journal of laboratory and clinical medicine**, Vol. 184, pp. 35-44.e4, (2017) ([PubMed](#)).

Wu, Gao, Ma, Fu, Zhang, Luo: "Characterisation of betatrophin concentrations in childhood and adolescent obesity and insulin resistance." in: **Pediatric diabetes**, Vol. 17, Issue 1, pp. 53-60, (2016) ([PubMed](#)).

There are more publications referencing this product on: [Product page](#)

Images



ELISA

Image 1. Typical standard curve