ANTIBODIES ONLINE

Datasheet for ABIN2345146 CEL ELISA Kit

2 Publications



Overview

Quantity:	96 tests
Target:	CEL
Reactivity:	Others
Method Type:	Competition ELISA
Application:	ELISA

Product Details

Purpose:	First, a CEL conjugate is coated on the ELISA plate. The unknown CEL protein samples or CEL-
	BSA standards are then added to the CEL conjugate preabsorbed plate. After a brief incubation,
	the anti-CEL monoclonal antibody is added, followed by an HRP conjugated secondary
	antibody. The content of CEL protein adducts in unknown samples is determined by
	comparison with the predetermined CEL-BSA standard curve. Despite the structure similarity
	between CEL and CML, the anti-CEL specific antibody in the OxiSelect™ CEL ELISA Kit will not
	cross react with CML protein adducts. 2
Brand:	OxiSelect™
Sample Type:	Serum, Plasma
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Components:	1. 96-well Protein Binding Plate : One strip well 96-well plate.
	2. Anti-CEL Antibody (1000X) : One 10 µL vial of anti-CEL antibody.
	3. Secondary Antibody, HRP Conjugate (1000Χ) : One 20 μL vial.
	4. Assay Diluent : One 50 mL bottle.

	5. 10X Wash Buffer : One 100 mL bottle.
	6. Substrate Solution : One 12 mL amber bottle.
	7. Stop Solution (Part. No. 310808): One 12 mL bottle.
	Box 2 (shipped on blue ice packs)
Material not included:	1. Protein samples such as purified protein, plasma, serum, cell lysate
	2. 1X PBS
	3. 10 μ L to 1000 μ L adjustable single channel micropipettes with disposable tips
	4. 50 μ L to 300 μ L adjustable multichannel micropipette with disposable tips
	5. Multichannel micropipette reservoir
	6. Microplate reader capable of reading at 450 nm (620 nm as optional reference wave length)
	3

Target Details

Target:	CEL
Background:	The non-enzymatic reaction of reducing carbohydrates with lysine side chains and N-terminal
	amino groups of macromolecules (proteins, phospholipids and nucleic acids) is called the
	Maillard reaction or glycation. The products of this process, termed advanced glycation end
	products (AGEs), adversely affect the functional properties of proteins, lipids and DNA. Tissue
	levels of AGE increase with age and the formation of AGEs is predominantly endogenous,
	though these products can also be derived from exogenous sources such as food and tobacco
	smoke. AGE modification of proteins can contribute to the pathophysiology of aging and long-
	term complications of diabetes, atherosclerosis and renal failure. AGEs also interact with a
	variety of cell-surface AGE-binding receptors (RAGE), leading either to their endocytosis and
	degradation or to cellular activation and pro-oxidant or pro- inflammatory events. $\boldsymbol{\epsilon}$ Although
	several AGE structures have been reported, it was demonstrated that N -(carboxymethyl) $\boldsymbol{\epsilon}$
	lysine (CML) and N -(carboxyethyl) lysine (CEL) are the major antigenic AGE structures. Next to
	glucose, reactive di-carbonyl compounds such as methylglyoxal are major precursors in the
	formation of cellular and extracellular AGEs. Methylglyoxal reacts with lysine residues to form
	CEL. CEL concentration is increased in patients who have diabetes with complications. $\boldsymbol{\epsilon}$ The
	quantity of CEL adduct in protein samples is determined by comparing its absorbance with that
	of a known CEL-BSA standard curve. Each kit provides sufficient reagents to perform up to 96
	assays, including standard curve and unknown protein samples.

Application Details

Comment:

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/4 | Product datasheet for ABIN2345146 | 07/26/2024 | Copyright antibodies-online. All rights reserved.

- Provides rapid detection and quantitation of CEL protein adducts
- · Perform up to 96 assays, including standard curve and unknown protein samples
- · Will not cross react with CML protein adducts

Plate:	Uncoated
Reagent Preparation:	CEL Conjugate Coated Plate: Note: The CEL Conjugate coated wells are not stable and should
	be used within 24 hrs after coating. Only coat the number of wells to be used immediately. 1.
	Immediately before use, prepare 1X Conjugate Diluent by diluting the 100X Conjugate Diluent in
	1X PBS. Example: Add 50 μL to 4.95 mL of 1X PBS. 2. Immediately before use, prepare 1.0 μ
	g/mL of CEL Conjugate by diluting the 1.0 mg/mL CEL Conjugate in 1X Conjugate Diluent.
	Example: Add 5 μ L of 1.0 mg/mL CEL Conjugate to 4.995 mL of 1X Conjugate Diluent and mix
	well. 3. Add 100 μ L of the 1 μ g/mL CEL Conjugate to each well to be tested and incubate
	overnight at 4 °C. Remove the CEL Conjugate coating solution and wash twice with 1X PBS.
	Blot plate on paper towels to remove excess fluid. Add 200 µL of Assay Diluent to each well and
	block for 1 hr at room temperature on an orbital shaker. Transfer the plate to 4 °C and remove
	the Assay Diluent immediately before use 1X Wash Buffer. Dilute the 10X Wash Buffer to 1X
	with deionized water. Stir to homogeneity. Anti-CEL Antibody and Secondary Antibody:
	Immediately before use dilute the Anti-CEL antibody 1:1000 and Secondary Antibody 1:1000
	with Assay Diluent. Do not store diluted solutions
	With Assay Dildent. Do not store dildted solutions.
Assay Procedure:	1. Prepare and mix all reagents thoroughly before use. Each CEL sample including unknown
	and standard should be assayed in duplicate.
	2. Add 50 µL of the unknown sample of CEL-BSA standard to the wells of the CEL conjugate
	before adding. Incubate at room temperature for 10 minutes on an orbital shaker.
	3. Add 50 µL of the diluted anti-CEL antibody to each well, incubate at room temperature for 1 hour on an orbital shaker.
	4. Wash 3 times with 250 μ L of 1X Wash Buffer with thorough aspiration between each wash.
	After the last wash, empty wells and tap microwell strips on absorbent pad or paper towel to remove excess 1X Wash Buffer.
	5. Add 100 μL of the diluted Secondary Antibody-HRP Conjugate to all wells and incubate for 1
	hour at room temperature on an orbital shaker. Wash the strip wells 3 times according to step 4 above.
	6. Warm Substrate Solution to room temperature. Add 100 L of Substrate Solution to each well.
	Incubate at room temperature for 5-20 minutes on an orbital shaker. Note: Watch plate
	carefully, if color changes rapidly, the reaction may need to be stopped sooner to prevent saturation.
	7. Stop the enzyme reaction by adding 100 μL of Stop Solution to each well. Results should be read immediately (color will fade over time).
	8. Read absorbance of each well on a microplate reader using 450 nm as the primary wave

Application Details	
	length. 5
Restrictions:	For Research Use only
Handling	
Handling Advice:	Avoid multiple freeze/thaw cycles.
Storage:	4 °C/-20 °C
Storage Comment:	Upon receipt, aliquot and store the Anti-CEL Antibody, CEL-BSA Standard, CEL Conjugate and 100X Conjugate Diluent at -20°C to avoid multiple freeze/thaw cycles. Store all other kit components at 4°C.
Publications	
Product cited in:	Cannizzaro, Rossoni, Savi, Altomare, Marinello, Saethang, Carini, Payne, Pisitkun, Aldini, Leelahavanichkul: "Regulatory landscape of AGE-RAGE-oxidative stress axis and its modulation by PPARγ activation in high fructose diet-induced metabolic syndrome." in: Nutrition & metabolism , Vol. 14, pp. 5, (2017) (PubMed).
	Morgan, Sheahan, Davies: "Perturbation of human coronary artery endothelial cell redox state and NADPH generation by methylglyoxal." in: PLoS ONE , Vol. 9, Issue 1, pp. e86564, (2014) (PubMed).