antibodies -online.com





Datasheet for ABIN2533976

GST ELISA Kit



Overview

96 tests GST Human 0.5 ng/mL - 30 ng/mL
Human
0.5 ng/mL - 30 ng/mL
0.5 ng/mL
ELISA
Human Glutathione S-Transferase ELISA Kit is an ELISA kit against Human Glutathione S-
Transferase.
Cell Culture Supernatant, Plasma, Serum
Quantitative
Colorimetric
0.1 ng/mL
Human Glutathione S-Transferases ELISA Kit is an ELISA kit against Human Glutathione S-
Transferases.
GST
Glutathione S-Transferases (GST Products)

Application Details

Application Notes:	Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less
	than 5 % within the expiration date under appropriate storage conditions. To minimize
	performance fluctuations, operation procedures and lab conditions should be strictly controlled.
	It is also strongly suggested that the whole assay is performed by the same user throughout.
	Recommended dilutions: Optimal dilutions/concentrations should be determined by the end
	user.
Comment:	The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5%
	within the expiration date under appropriate storage conditions. To minimize performance
	fluctuations, operation procedures and lab conditions should be strictly controlled. It is also
	strongly suggested that the whole assay is performed by the same user throughout.
Plate:	Pre-coated
Restrictions:	For Research Use only
Handling	
Storage:	4 °C/-20 °C
otorage.	7 0/ 20 0
Storage Comment:	Upon receipt, store the kit according to the storage instruction in the kit's manual.
Expiry Date:	6 months