# antibodies -online.com





# Datasheet for ABIN2534617

## **HYAL3 ELISA Kit**



#### Overview

| Quantity:                | 96 tests            |
|--------------------------|---------------------|
| Target:                  | HYAL3               |
| Reactivity:              | Human               |
| Detection Range:         | 0.1 ng/mL - 8 ng/mL |
| Minimum Detection Limit: | 0.1 ng/mL           |
| Application:             | ELISA               |

### **Product Details**

| Purpose:           | Human Hyaluronidase 3 ELISA Kit is an ELISA kit against Human Hyaluronidase 3. |
|--------------------|--|
| Sample Type:       | Cell Culture Supernatant, Plasma, Serum  |
| Analytical Method: | Quantitative   |
| Detection Method:  | Colorimetric   |
| Sensitivity:       | 0.01 ng/mL   |
| Characteristics:   | Human Hyaluronidase 3 ELISA Kit is an ELISA kit against Human Hyaluronidase 3. |

# Target Details

| Target:           | HYAL3                               |
|-------------------|-------------------------------------|
| Alternative Name: | Hyaluronidase 3 (HYAL3 Products)    |
| Pathways:         | Glycosaminoglycan Metabolic Process |

# **Application Details**

| Application Notes: | Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less |
|--------------------|---|
|                    | than 5 $\%$ within the expiration date under appropriate storage conditions. To minimize              |
|                    | performance fluctuations, operation procedures and lab conditions should be strictly controlled.      |
|                    | It is also strongly suggested that the whole assay is performed by the same user throughout.          |
|                    | Recommended dilutions: Optimal dilutions/concentrations should be determined by the end               |
|                    | user.   |
| Comment:           | The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5%    |
|                    | within the expiration date under appropriate storage conditions. To minimize performance              |
|                    | fluctuations, operation procedures and lab conditions should be strictly controlled. It is also       |
|                    | strongly suggested that the whole assay is performed by the same user throughout.                     |
| Plate:             | Pre-coated  |
| Restrictions:      | For Research Use only   |
| Handling           |   |
| Storage:           | 4 °C/-20 °C   |
| Storage Comment:   | Upon receipt, store the kit according to the storage instruction in the kit's manual.                 |
| Expiry Date:       | 6 months  |