

Datasheet for ABIN2537009 NQ01 ELISA Kit



[Go to Product page](#)

Overview

Quantity: 96 tests

Target: NQ01

Reactivity: Rat

Method Type: Sandwich ELISA

Detection Range: 0.312 ng/mL - 20 ng/mL

Minimum Detection Limit: 0.312 ng/mL

Application: ELISA

Product Details

Purpose: Rat NQ01 ELISA Kit is a sandwich ELISA kit for use with Tissue homogenates, cell lysates and other biological fluids. This assay has high sensitivity and excellent specificity for detection of NQ01 (NQ01)

No significant cross-reactivity or interference between NQ01 (NQ01) and analogues was observed.

Sample Type: Cell Lysate, Tissue Homogenate

Analytical Method: Quantitative

Detection Method: Colorimetric

Specificity: This assay has high sensitivity and excellent specificity for detection of NADH Dehydrogenase, Quinone 1 (NQ01)

Sensitivity: < 0.115 ng/mL

Target Details

Target:	NQ01
Alternative Name:	NADH Dehydrogenase, Quinone 1 (NQ01) (NQ01 Products)

Application Details

Application Notes:	<p>Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout.</p> <p>Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user.</p> <p>Standard Form: Lyophilized</p>
Comment:	<p>The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within the expiration date under appropriate storage conditions. - minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout.</p>
Plate:	Pre-coated
Restrictions:	For Research Use only

Handling

Storage:	4 °C/-20 °C
Storage Comment:	Upon receipt, store the kit according to the storage instruction in the kit's manual.
Expiry Date:	6 months