

## Datasheet for ABIN2537725

## **PER1 ELISA Kit**



## Overview

Quantity:	96 tests
Target:	PER1
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.312 ng/mL - 20 ng/mL
Minimum Detection Limit:	0.312 ng/mL
Application:	ELISA
Product Details	
Purpose:	Human PER1 ELISA Kit is a sandwich ELISA kit for use with Tissue homogenates and other
	biological fluids. This assay has high sensitivity and excellent specificity for detection of Period
	Circadian Protein 1 (PER1)
	No significant cross-reactivity or interference between Period Circadian Protein 1 (PER1) and
	analogues was observed.
Sample Type:	Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Period Circadian Protein
	1 (PER1)
Sensitivity:	< 0.122 ng/mL

## **Target Details**

Target:	PER1
Alternative Name:	Period Circadian Protein 1 (PER1) (PER1 Products)
Pathways:	Photoperiodism
Application Details	
Application Notes:	Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled It is also strongly suggested that the whole assay is performed by the same user throughout. Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user.  Standard Form: Lyophilized
Comment:	The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within the expiration date under appropriate storage conditions minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout.
Plate:	Pre-coated
Restrictions:	For Research Use only
Handling	
Storage:	4 °C/-20 °C
Storage Comment:	Upon receipt, store the kit according to the storage instruction in the kit's manual.
Expiry Date:	6 months