

Datasheet for ABIN2538240

PCSK2 ELISA Kit



Overview

| Quantity: | 96 tests |
|--------------------------|--|
| Target: | PCSK2 |
| Reactivity: | Human |
| Method Type: | Sandwich ELISA |
| Detection Range: | 0.312 ng/mL - 20 ng/mL |
| Minimum Detection Limit: | 0.312 ng/mL |
| Application: | ELISA |
| Product Details | |
| Purpose: | Human PCSK2 ELISA Kit is a sandwich ELISA kit for use with Tissue homogenates and other |
| | biological fluids. This assay has high sensitivity and excellent specificity for detection of |
| | Proprotein Convertase Subtilisin/Kexin Type 2 (PCSK2) |
| | No significant cross-reactivity or interference between Proprotein Convertase Subtilisin/Kexin |
| | Type 2 (PCSK2) and analogues was observed. |
| Sample Type: | Tissue Homogenate |
| Analytical Method: | Quantitative |
| Detection Method: | Colorimetric |
| Specificity: | This assay has high sensitivity and excellent specificity for detection of Proprotein Convertase |
| | Subtilisin/Kexin Type 2 (PCSK2) |
| Sensitivity: | < 0.112 ng/mL |

Target Details

| rarget betans | |
|---------------------|---|
| Target: | PCSK2 |
| Abstract: | PCSK2 Products |
| Pathways: | Peptide Hormone Metabolism, cAMP Metabolic Process, Maintenance of Protein Location, Negative Regulation of Transporter Activity |
| Application Details | |
| Application Notes: | Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout. Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user. Standard Form: Lyophilized |
| Comment: | The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within the expiration date under appropriate storage conditions minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout. |
| Plate: | Pre-coated |
| Restrictions: | For Research Use only |
| Handling | |
| Storage: | 4 °C/-20 °C |
| Storage Comment: | Upon receipt, store the kit according to the storage instruction in the kit's manual. |
| Expiry Date: | 6 months |