

Datasheet for ABIN2538671 **RNASE8 ELISA Kit**



[Go to Product page](#)

Overview

Quantity: 96 tests

Target: RNASE8

Reactivity: Human

Method Type: Sandwich ELISA

Detection Range: 0.312 ng/mL - 20 ng/mL

Minimum Detection Limit: 0.312 ng/mL

Application: ELISA

Product Details

Purpose: Human Ribonuclease A8 ELISA Kit is a sandwich ELISA kit for use with Tissue homogenates and other biological fluids. This assay has high sensitivity and excellent specificity for detection of Ribonuclease A8 (RNASE8)

No significant cross-reactivity or interference between Ribonuclease A8 (RNASE8) and analogues was observed.

Sample Type: Tissue Homogenate

Analytical Method: Quantitative

Detection Method: Colorimetric

Specificity: This assay has high sensitivity and excellent specificity for detection of Ribonuclease A8 (RNASE8)

Sensitivity: < 0.109 ng/mL

Target Details

Target:	RNASE8
Alternative Name:	Ribonuclease A8 (RNASE8) (RNASE8 Products)

Application Details

Application Notes:	<p>Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout.</p> <p>Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user.</p> <p>Standard Form: Lyophilized</p>
Comment:	<p>The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within the expiration date under appropriate storage conditions. - minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout.</p>
Plate:	Pre-coated
Restrictions:	For Research Use only

Handling

Storage:	4 °C/-20 °C
Storage Comment:	Upon receipt, store the kit according to the storage instruction in the kit's manual.
Expiry Date:	6 months