

## Datasheet for ABIN2538934 SAA4 ELISA Kit



Overview

Quantity:	96 tests
Target:	SAA4
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.156 ng/mL - 10 ng/mL
Minimum Detection Limit:	0.156 ng/mL
Application:	ELISA
Product Details	
Purpose:	- Human SAA4 ELISA Kit is a sandwich ELISA kit for use with Serum, plasma and other biological
	fluids. This assay has high sensitivity and excellent specificity for detection of Serum Amyloid

No significant cross-reactivity or interference between Serum Amyloid A4, Constitutive (SAA4) and analogues was observed.

	-
Sample Type:	Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Serum Amyloid A4, Constitutive (SAA4)
Sensitivity:	< 0.063 ng/mL

A4, Constitutive (SAA4)

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/2 | Product datasheet for ABIN2538934 | 07/26/2024 | Copyright antibodies-online. All rights reserved.

Target:	SAA4
Alternative Name:	Serum Amyloid A4, Constitutive (SAA4) (SAA4 Products)
Application Details	
Application Notes:	Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less
	than 5 % within the expiration date under appropriate storage conditions. To minimize
	performance fluctuations, operation procedures and lab conditions should be strictly controlled
	It is also strongly suggested that the whole assay is performed by the same user throughout.
	Recommended dilutions: Optimal dilutions/concentrations should be determined by the end
	user.
	Standard Form: Lyophilized
Comment:	The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5%
	within the expiration date under appropriate storage conditions minimize performance
	fluctuations, operation procedures and lab conditions should be strictly controlled. It is also
	strongly suggested that the whole assay is performed by the same user throughout.

Expiry Date:

Storage Comment:

Restrictions:

Handling

Storage:

Upon receipt, store the kit according to the storage instruction in the kit's manual. 6 months

4 °C/-20 °C

For Research Use only