

## Datasheet for ABIN2540232 Urocortin 2 ELISA Kit



Overview

Quantity:	96 tests
Target:	Urocortin 2 (UCN2)
Reactivity:	Mouse
Method Type:	Competition ELISA
Detection Range:	1.23 ng/mL - 100 ng/mL
Minimum Detection Limit:	1.23 ng/mL
Application:	ELISA
Product Details	
Purpose:	Mouse Urocortin 2 ELISA Kit is a competitive ELISA kit for use with Serum, plasma and other biological fluids. This assay has high sensitivity and excellent specificity for detection of Urocortin 2 (UCN2) No significant cross-reactivity or interference between Urocortin 2 (UCN2) and analogues was observed.
Purpose: Sample Type:	biological fluids. This assay has high sensitivity and excellent specificity for detection of Urocortin 2 (UCN2) No significant cross-reactivity or interference between Urocortin 2 (UCN2) and analogues was
	biological fluids. This assay has high sensitivity and excellent specificity for detection of Urocortin 2 (UCN2) No significant cross-reactivity or interference between Urocortin 2 (UCN2) and analogues was observed.
Sample Type:	biological fluids. This assay has high sensitivity and excellent specificity for detection of Urocortin 2 (UCN2) No significant cross-reactivity or interference between Urocortin 2 (UCN2) and analogues was observed. Plasma, Serum
Sample Type: Analytical Method:	biological fluids. This assay has high sensitivity and excellent specificity for detection of Urocortin 2 (UCN2) No significant cross-reactivity or interference between Urocortin 2 (UCN2) and analogues was observed. Plasma, Serum Quantitative

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/2 | Product datasheet for ABIN2540232 | 07/26/2024 | Copyright antibodies-online. All rights reserved.

Target Details	
Target:	Urocortin 2 (UCN2)
Abstract:	UCN2 Products
UniProt:	Q99ML8
Pathways:	Positive Regulation of Peptide Hormone Secretion, Negative Regulation of Hormone Secretion, cAMP Metabolic Process
Application Details	
Application Notes:	<ul> <li>Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less</li> <li>than 5 % within the expiration date under appropriate storage conditions. To minimize</li> <li>performance fluctuations, operation procedures and lab conditions should be strictly controlled.</li> <li>It is also strongly suggested that the whole assay is performed by the same user throughout.</li> <li>Recommended dilutions: Optimal dilutions/concentrations should be determined by the end</li> <li>user.</li> <li>Standard Form: Lyophilized</li> </ul>
Comment:	The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within the expiration date under appropriate storage conditions minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout.
Plate:	Pre-coated
Restrictions:	For Research Use only
Handling	
Storage:	4 °C/-20 °C
Storage Comment:	Upon receipt, store the kit according to the storage instruction in the kit's manual.
Expiry Date:	6 months