

Datasheet for ABIN2540842

AOAH ELISA Kit



Overview

| Quantity: | 96 tests |
|--------------------------|---|
| Target: | AOAH |
| Reactivity: | Human |
| Method Type: | Sandwich ELISA |
| Detection Range: | 0.156 ng/mL - 10 ng/mL |
| Minimum Detection Limit: | 0.156 ng/mL |
| Application: | ELISA |
| Product Details | |
| Purpose: | Human Acyloxyacyl Hydrolase ELISA Kit is a sandwich ELISA kit for use with serum, plasma, |
| | tissue homogenates, cell lysates and other biological fluids. This assay has high sensitivity and |
| | excellent specificity for detection of Acyloxyacyl Hydrolase (AOAH) |
| | No significant cross-reactivity or interference between Acyloxyacyl Hydrolase (AOAH) and |
| | analogues was observed. |
| Sample Type: | Cell Lysate, Plasma, Serum, Tissue Homogenate |
| Analytical Method: | Quantitative |
| Detection Method: | Colorimetric |
| Specificity: | This assay has high sensitivity and excellent specificity for detection of Acyloxyacyl Hydrolase |
| | (AOAH) |
| Sensitivity: | < 0.058 ng/mL |

Target Details

| Application Details Application Notes: Stability: The stability of the kit is determined by the rate of activity loss. The loss of than 5 % within the expiration date under appropriate storage conditions. To minimperformance fluctuations, operation procedures and lab conditions should be strictly lit is also strongly suggested that the whole assay is performed by the same user the Recommended dilutions: Optimal dilutions/concentrations should be determined user. Standard Form: Lyophilized | |
|--|------------------|
| Application Notes: Stability: The stability of the kit is determined by the rate of activity loss. The loss of than 5 % within the expiration date under appropriate storage conditions. To minimperformance fluctuations, operation procedures and lab conditions should be strictly lit is also strongly suggested that the whole assay is performed by the same user to Recommended dilutions: Optimal dilutions/concentrations should be determined user. | |
| than 5 % within the expiration date under appropriate storage conditions. To minimperformance fluctuations, operation procedures and lab conditions should be strictly likely also strongly suggested that the whole assay is performed by the same user to Recommended dilutions: Optimal dilutions/concentrations should be determined user. | |
| performance fluctuations, operation procedures and lab conditions should be strictly like also strongly suggested that the whole assay is performed by the same user to Recommended dilutions: Optimal dilutions/concentrations should be determined user. | mize |
| It is also strongly suggested that the whole assay is performed by the same user to Recommended dilutions: Optimal dilutions/concentrations should be determined user. | |
| Recommended dilutions: Optimal dilutions/concentrations should be determined user. | ctly controlled. |
| user. | throughout. |
| | by the end |
| Standard Form: Lyophilizad | |
| Standard Form. Lyophilized | |
| Comment: The stability of the kit is determined by the rate of activity loss. The loss rate is less | s than 5% |
| within the expiration date under appropriate storage conditions minimize perform | mance |
| fluctuations, operation procedures and lab conditions should be strictly controlled | l. It is also |
| strongly suggested that the whole assay is performed by the same user throughout | ut. |
| Plate: Pre-coated | |
| Restrictions: For Research Use only | |
| | |
| Handling | |
| Storage: 4 °C/-20 °C | |
| Storage Comment: Upon receipt, store the kit according to the storage instruction in the kit's manual. | |
| Expiry Date: 6 months | |