

Datasheet for ABIN2542220 BIN1 ELISA Kit



Overview

| 010111011 | |
|--------------------------|--|
| Quantity: | 96 tests |
| Target: | BIN1 |
| Reactivity: | Human |
| Method Type: | Sandwich ELISA |
| Detection Range: | 0.156 ng/mL - 10 ng/mL |
| Minimum Detection Limit: | 0.156 ng/mL |
| Application: | ELISA |
| Product Details | |
| Purpose: | Human Bridging Integrator 1 ELISA Kit is a sandwich ELISA kit for use with Serum, plasma and other biological fluids. This assay has high sensitivity and excellent specificity for detection of Bridging Integrator 1 (BIN1) No significant cross-reactivity or interference between Bridging Integrator 1 (BIN1) and analogues was observed. |
| Sample Type: | Plasma, Serum |
| Analytical Method: | Quantitative |
| Detection Method: | Colorimetric |
| Specificity: | This assay has high sensitivity and excellent specificity for detection of Bridging Integrator 1 |

Sensitivity:

< 0.061 ng/mL

(BIN1)

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/2 | Product datasheet for ABIN2542220 | 07/26/2024 | Copyright antibodies-online. All rights reserved.

| Target Details | |
|---------------------|--|
| Target: | BIN1 |
| Abstract: | BIN1 Products |
| Application Details | |
| Application Notes: | Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout. Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user. Standard Form: Lyophilized |
| Comment: | The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within the expiration date under appropriate storage conditions minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout. |
| Plate: | Pre-coated |
| Restrictions: | For Research Use only |
| Handling | |
| Storage: | 4 °C/-20 °C |
| Storage Comment: | Upon receipt, store the kit according to the storage instruction in the kit's manual. |
| Expiry Date: | 6 months |