

Datasheet for ABIN2542898

CCK8 ELISA Kit



Overview

Quantity:	96 tests
Target:	CCK8
Reactivity:	Human
Method Type:	Competition ELISA
Detection Range:	12.35 pg/mL - 1000 pg/mL
Minimum Detection Limit:	12.35 pg/mL
Application:	ELISA

Product Details	
Purpose:	Human CCK8 ELISA Kit is a competitive ELISA kit for use with Serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids. This assay has high sensitivity and excellent specificity for detection of Cholecystokinin 8, Octapeptide (CCK8) No significant cross-reactivity or interference between Cholecystokinin 8, Octapeptide (CCK8) and analogues was observed.
Sample Type:	Cell Culture Supernatant, Cell Lysate, Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Cholecystokinin 8, Octapeptide (CCK8)
Sensitivity:	< 4.63 pg/mL

Target Details

Target:	CCK8
Abstract:	CCK8 Products
Application Details	
Application Notes:	Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout. Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user. Standard Form: Lyophilized
Comment:	The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout.
Plate:	Pre-coated
Restrictions:	For Research Use only
Handling	
Storage:	4 °C/-20 °C
Storage Comment:	Upon receipt, store the kit according to the storage instruction in the kit's manual.
Expiry Date:	6 months