# antibodies - online.com







# **OB Cadherin ELISA Kit**



Image



### Overview

Quantity:	96 tests
Target:	OB Cadherin (CDH11)
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.8-200 ng/mL
Minimum Detection Limit:	0.8 ng/mL
Application:	ELISA

Product Details	
Purpose:	Human Cadherin-11 ELISA Kit for Cell Culture Supernatants, Plasma, and Serum samples.
Sample Type:	Cell Culture Supernatant, Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	The antibody pair provided in this kit recognizes Human Cadherin-11 (CDH11)
Sensitivity:	0.8 ng/mL
Characteristics:	<ul> <li>Strip plates and additional reagents allow for use in multiple experiments</li> <li>Quantitative protein detection</li> <li>Establishes normal range</li> <li>The best products for confirmation of antibody array data</li> </ul>
Components:	Pre-Coated 96-well Strip Microplate

### **Product Details**

- · Wash Buffer
- · Stop Solution
- · Assay Diluent(s)
- · Lyophilized Standard
- · Biotinylated Detection Antibody
- · Streptavidin-Conjugated HRP
- · TMB One-Step Substrate

### Material not included:

- · Distilled or deionized water
- Precision pipettes to deliver 2  $\mu$ L to 1  $\mu$ L volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 µL and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- · Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- · Log-log graph paper or computer and software for ELISA data analysis

### **Target Details**

Target:	OB Cadherin (CDH11)
Alternative Name:	CDH-11 (CDH11 Products)
Background:	Gene Names: CDH11  Protein names: Cadherin-11 (OSF-4) (Osteoblast cadherin) (OB-cadherin)
Gene ID:	1009

### **Application Details**

Application Notes:	Recommended Dilution for serum and plasma samples2 fold
Sample Volume:	100 μL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual.
	2. Add 100 μL of standard or sample to each well.
	3. Incubate 2.5 h at RT or O/N at 4 °C.
	4. Add 100 μL of prepared biotin antibody to each well.
	5. Incubate 1 h at RT.
	6. Add 100 $\mu L$ of prepared Streptavidin solution to each well.
	7. Incubate 45 min at RT.
	7. Incubate 45 min at KT.

8. Add 100  $\mu L$  of TMB One-Step Substrate Reagent to each well.

9. Incubate 30 min at RT.

10. Add 50 µL of Stop Solution to each well.

11. Read at 450 nm immediately.

Restrictions:

For Research Use only

# Handling

Storage:	-20 °C
Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated
	freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is
	recommended to store at -80°C.
Expiry Date:	6 months

## **Images**

# Assay Diluent D

Human CDH-11 concentration (ng/ml)

### **ELISA**

Image 1.