antibodies .- online.com





Datasheet for ABIN2702942

CXCL9 ELISA Kit





Overview

Quantity:	96 tests
Target:	CXCL9
Reactivity:	Pig
Method Type:	Sandwich ELISA
Detection Range:	5-1.400 pg/mL
Minimum Detection Limit:	5 pg/mL
Application:	ELISA

Product Details

Purpose:	Porcine MIG (CXCL9) ELISA Kit for cell culture supernatants, plasma, and serum samples.
Sample Type:	Cell Culture Supernatant, Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	Cross Reactivity: This antibody pair shows no cross-reactivity with human, mouse, bovine and feline CXCL9/MIG
Sensitivity:	5 pg/mL
Characteristics:	 Strip plates and additional reagents allow for use in multiple experiments Quantitative protein detection Establishes normal range The best products for confirmation of antibody array data

Product Details

Components:

- · Pre-Coated 96-well Strip Microplate
- · Wash Buffer
- · Stop Solution
- · Assay Diluent(s)
- · Lyophilized Standard
- · Biotinylated Detection Antibody
- · Streptavidin-Conjugated HRP
- TMB One-Step Substrate

Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2 μL to 1 μL volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 µL and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- · Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target:	CXCL9
Alternative Name:	CXCL9 (CXCL9 Products)
Background:	Gene Names: CXCL9 Protein names: C-X-C motif chemokine
Gene ID:	100135681
UniProt:	B0FYK2

Application Details

Application Notes:	Recommended Dilution for serum and plasma samples2 fold
Sample Volume:	100 μL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual.
	2. Add 100 μL of standard or sample to each well.
	3. Incubate 2.5 h at RT or O/N at 4 °C.
	4. Add 100 μL of prepared biotin antibody to each well.
	5. Incubate 1 h at RT.
	6. Add 100 μL of prepared Streptavidin solution to each well.

7. Incubate 45 min at RT.

- 8. Add 100 μL of TMB One-Step Substrate Reagent to each well.
- 9. Incubate 30 min at RT.
- 10. Add 50 μ L of Stop Solution to each well.
- 11. Read at 450 nm immediately.


Restrictions:

For Research Use only

Handling

Storage:	-20 °C
Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated
	freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is
	recommended to store at -80°C.
Expiry Date:	6 months

Images



ELISA

Image 1.