# antibodies -online.com





Datasheet for ABIN2702960

# **CD226 ELISA Kit**





_					
U	V	er	VI	е	W

Quantity:	96 tests
Target:	CD226
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	5-600 ng/mL
Minimum Detection Limit:	5 ng/mL
Application:	ELISA

Product Details	
Purpose:	Human DNAM-1 (CD226) ELISA Kit for cell culture supernatants, plasma, and serum samples.
Sample Type:	Cell Culture Supernatant, Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA antibody pair detects human DNAM1. Other species not determined.
Sensitivity:	5 ng/mL
Characteristics:	<ul> <li>Strip plates and additional reagents allow for use in multiple experiments</li> <li>Quantitative protein detection</li> <li>Establishes normal range</li> <li>The best products for confirmation of antibody array data</li> </ul>
Components:	Pre-Coated 96-well Strip Microplate

#### **Product Details**

- · Wash Buffer
- · Stop Solution
- · Assay Diluent(s)
- · Lyophilized Standard
- · Biotinylated Detection Antibody
- · Streptavidin-Conjugated HRP
- · TMB One-Step Substrate

#### Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2  $\mu$ L to 1  $\mu$ L volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 µL and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- · Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- · Log-log graph paper or computer and software for ELISA data analysis

#### **Target Details**

Target:	CD226
Alternative Name:	DNAM-1 (CD226 Products)
Background:	Gene Names: CD226 DNAM1  Protein names: CD226 antigen (DNAX accessory molecule 1) (DNAM-1) (CD antigen CD226)
Gene ID:	10666
UniProt:	Q15762
Pathways:	Regulation of Leukocyte Mediated Immunity, Positive Regulation of Immune Effector Process, Cancer Immune Checkpoints

#### **Application Details**

Application Notes:	Recommended Dilution for serum and plasma samples2-20 fold
Sample Volume:	100 μL
Plate:	Pre-coated
Protocol:	<ol> <li>Prepare all reagents, samples and standards as instructed in the manual.</li> <li>Add 100 µL of standard or sample to each well.</li> <li>Incubate 2.5 h at RT or O/N at 4 °C.</li> <li>Add 100 µL of prepared biotin antibody to each well.</li> </ol>

5. Incubate 1 h at RT.

- 6. Add 100  $\mu L$  of prepared Streptavidin solution to each well.
- 7. Incubate 45 min at RT.
- 8. Add 100  $\mu L$  of TMB One-Step Substrate Reagent to each well.
- 9. Incubate 30 min at RT.
- 10. Add 50 µL of Stop Solution to each well.
- 11. Read at 450 nm immediately.

Restrictions:


For Research Use only

## Handling

Storage:	-20 °C
Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated
	freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is
	recommended to store at -80°C.

## **Images**

Expiry Date:



6 months

#### **ELISA**

Image 1.