

Datasheet for ABIN2703012
FGF23 ELISA Kit



[Go to Product page](#)

1 Image

Overview

Quantity:	96 tests
Target:	FGF23
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.3-75 ng/mL
Minimum Detection Limit:	0.3 ng/mL
Application:	ELISA

Product Details

Purpose:	Human FGF-23 ELISA Kit for Cell Culture Supernatants, Plasma, and Serum samples.
Sample Type:	Cell Culture Supernatant, Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	The antibody pair provided in this kit recognizes Human FGF-23
Sensitivity:	0.3 ng/mL
Characteristics:	<ul style="list-style-type: none">• Strip plates and additional reagents allow for use in multiple experiments• Quantitative protein detection• Establishes normal range• The best products for confirmation of antibody array data
Components:	<ul style="list-style-type: none">• Pre-Coated 96-well Strip Microplate

Product Details

- Wash Buffer
- Stop Solution
- Assay Diluent(s)
- Lyophilized Standard
- Biotinylated Detection Antibody
- Streptavidin-Conjugated HRP
- TMB One-Step Substrate

Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2 µL to 1 µL volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 µL and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target: FGF23

Alternative Name: FGF-23 ([FGF23 Products](#))

Background: Gene Names: FGF23 HYPF UNQ3027/PRO9828
Protein names: Fibroblast growth factor 23 (FGF-23) (Phosphatonin) (Tumor-derived hypophosphatemia-inducing factor) [Cleaved into: Fibroblast growth factor 23 N-terminal peptide, Fibroblast growth factor 23 C-terminal peptide]

Gene ID: 8074

UniProt: [Q9GZV9](#)

Pathways: [RTK Signaling](#), [Fc-epsilon Receptor Signaling Pathway](#), [EGFR Signaling Pathway](#), [Neurotrophin Signaling Pathway](#), [Negative Regulation of Hormone Secretion](#)

Application Details

Application Notes: Recommended Dilution for serum and plasma samples 2 fold

Sample Volume: 100 µL

Plate: Pre-coated

Protocol: 1. Prepare all reagents, samples and standards as instructed in the manual.

Application Details

2. Add 100 μ L of standard or sample to each well.
3. Incubate 2.5 h at RT or O/N at 4 $^{\circ}$ C.
4. Add 100 μ L of prepared biotin antibody to each well.
5. Incubate 1 h at RT.
6. Add 100 μ L of prepared Streptavidin solution to each well.
7. Incubate 45 min at RT.
8. Add 100 μ L of TMB One-Step Substrate Reagent to each well.
9. Incubate 30 min at RT.
10. Add 50 μ L of Stop Solution to each well.
11. Read at 450 nm immediately.

Restrictions: For Research Use only


Handling

Storage: -20 $^{\circ}$ C

Storage Comment: The entire kit may be stored at -20 $^{\circ}$ C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4 $^{\circ}$ C for up to 6 months. For extended storage, it is recommended to store at -80 $^{\circ}$ C.

Expiry Date: 6 months

Images



ELISA

Image 1.