



[Go to Product page](#)

Datasheet for ABIN2703076

HCG beta ELISA Kit

1 Image

Overview

Quantity:	96 tests
Target:	HCG beta
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	40-10.000 pg/mL
Minimum Detection Limit:	40 pg/mL
Application:	ELISA

Product Details

Purpose:	Human hCG beta ELISA Kit for cell culture supernatants, plasma, and serum samples.
Sample Type:	Cell Culture Supernatant, Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA antibody pair detects human hCG-beta. Other species not determined.
Sensitivity:	40 pg/mL
Characteristics:	<ul style="list-style-type: none">• Strip plates and additional reagents allow for use in multiple experiments• Quantitative protein detection• Establishes normal range• The best products for confirmation of antibody array data
Components:	<ul style="list-style-type: none">• Pre-Coated 96-well Strip Microplate

Product Details

- Wash Buffer
- Stop Solution
- Assay Diluent(s)
- Lyophilized Standard
- Biotinylated Detection Antibody
- Streptavidin-Conjugated HRP
- TMB One-Step Substrate

Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2 μ L to 1 μ L volumes
- Adjustable 1-25 μ L pipettes for reagent preparation
- 100 μ L and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target: HCG beta

Alternative Name: hCG-beta ([HCG beta Products](#))

Background: Gene Names: CGB CGB3, CGB5, CGB7, CGB8
Protein names: Choriongonadotropin subunit beta (CG-beta) (Chorionic gonadotrophin chain beta)

Gene ID: 1082, 93659, 94115

UniProt: [P0DN86](#)

Application Details

Application Notes: Recommended Dilution for serum and plasma samples 2-20 fold

Sample Volume: 100 μ L

Plate: Pre-coated

Protocol:

1. Prepare all reagents, samples and standards as instructed in the manual.
2. Add 100 μ L of standard or sample to each well.
3. Incubate 2.5 h at RT or O/N at 4 $^{\circ}$ C.
4. Add 100 μ L of prepared biotin antibody to each well.
5. Incubate 1 h at RT.

Application Details

6. Add 100 μ L of prepared Streptavidin solution to each well.
7. Incubate 45 min at RT.
8. Add 100 μ L of TMB One-Step Substrate Reagent to each well.
9. Incubate 30 min at RT.
10. Add 50 μ L of Stop Solution to each well.
11. Read at 450 nm immediately.

Restrictions: For Research Use only


Handling

Storage: -20 $^{\circ}$ C

Storage Comment: The entire kit may be stored at -20 $^{\circ}$ C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4 $^{\circ}$ C for up to 6 months. For extended storage, it is recommended to store at -80 $^{\circ}$ C.

Expiry Date: 6 months

Images



ELISA

Image 1.