

Datasheet for ABIN2703452

SERPINA4 ELISA Kit





Go to Product page

_				
()	ve.	rv/	101	Λ

Quantity:	96 tests
Target:	SERPINA4
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.12-35 ng/mL
Minimum Detection Limit:	0.12 ng/mL
Application:	ELISA

Product Details

Product Details		
Purpose:	Human SERPINA4 ELISA Kit for cell culture supernatants, plasma, and serum samples.	
Sample Type:	Plasma, Cell Culture Supernatant, Serum	
Analytical Method:	Quantitative	
Detection Method:	Colorimetric	
Specificity:	This ELISA antibody pair detects human Serpin A4. Other species not determined.	
Sensitivity:	0.12 ng/mL	
Characteristics:	 Strip plates and additional reagents allow for use in multiple experiments Quantitative protein detection Establishes normal range The best products for confirmation of antibody array data 	
Components:	Pre-Coated 96-well Strip Microplate	

Product Details

- · Wash Buffer
- · Stop Solution
- · Assay Diluent(s)
- · Lyophilized Standard
- · Biotinylated Detection Antibody
- · Streptavidin-Conjugated HRP
- · TMB One-Step Substrate

Material not included:

- · Distilled or deionized water
- Precision pipettes to deliver 2 μ L to 1 μ L volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 μL and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- · Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- · Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target:	SERPINA4
Alternative Name:	Serpin A4 (SERPINA4 Products)
Background:	Gene Names: SERPINA4 KST PI4 Protein names: Kallistatin (Kallikrein inhibitor) (Peptidase inhibitor 4) (PI-4) (Serpin A4)
Gene ID:	5267
UniProt:	P29622

Application Details

Recommended Dilution for serum and plasma samples10,000 fold	
100 μL	
Pre-coated	
1. Prepare all reagents, samples and standards as instructed in the manual.	
2. Add 100 µL of standard or sample to each well.	
3. Incubate 2.5 h at RT or O/N at 4 °C.	
4. Add 100 μ L of prepared biotin antibody to each well.	
5. Incubate 1 h at RT.	
6. Add 100 μL of prepared Streptavidin solution to each well.	
7. Incubate 45 min at RT.	

8. Add 100 μL of TMB One-Step Substrate Reagent to each well.

9. Incubate 30 min at RT.

10. Add 50 µL of Stop Solution to each well.

11. Read at 450 nm immediately.

Restrictions:

For Research Use only

Handling

Storage:	-20 °C
Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated
	freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is
	recommended to store at -80°C.


Expiry Date:

6 months

Images

Assay Diluent A OD=450 nm 0.01





ELISA

Image 1.