antibodies - online.com





Datasheet for ABIN2703454

alpha 2 Antiplasmin ELISA Kit





Characteristics:

Overview	
Quantity:	96 tests
Target:	alpha 2 Antiplasmin (SERPINF2)
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.12-30 ng/mL
Minimum Detection Limit:	0.12 ng/mL
Application:	ELISA
Product Details	
Purpose:	Human SERPINF2 (Alpha-2-Antiplasmin) ELISA Kit for Cell Culture Supernatants, Plasma, and Serum samples.
Sample Type:	Cell Culture Supernatant, Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA antibody pair recognizes human SERPINF2. Other species not yet determined.
Sensitivity:	0.12 ng/mL

· The best products for confirmation of antibody array data

· Quantitative protein detection • Establishes normal range

• Strip plates and additional reagents allow for use in multiple experiments

Product Details

Components:

- · Pre-Coated 96-well Strip Microplate
- · Wash Buffer
- · Stop Solution
- · Assay Diluent(s)
- · Lyophilized Standard
- · Biotinylated Detection Antibody
- · Streptavidin-Conjugated HRP
- TMB One-Step Substrate

Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2 μL to 1 μL volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 µL and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- · Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- · Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target:	alpha 2 Antiplasmin (SERPINF2)
Alternative Name:	Serpin F2 (SERPINF2 Products)
Background:	Gene Names: SERPINF2 AAP PLI Protein names: Alpha-2-antiplasmin (Alpha-2-AP) (Alpha-2-plasmin inhibitor) (Alpha-2-Pl) (Serpin F2)
Gene ID:	5345
UniProt:	P08697
Pathways:	Regulation of Systemic Arterial Blood Pressure by Hormones

Application Details

Application Notes:	Recommended Dilution for serum and plasma samples200 - 2,000 fold
Sample Volume:	100 μL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual. 2. Add 100 μ L of standard or sample to each well. 3. Incubate 2.5 h at RT or O/N at 4 °C.

- 4. Add 100 μL of prepared biotin antibody to each well.
- 5. Incubate 1 h at RT.
- 6. Add 100 µL of prepared Streptavidin solution to each well.
- 7. Incubate 45 min at RT.
- 8. Add 100 µL of TMB One-Step Substrate Reagent to each well.
- 9. Incubate 30 min at RT.
- 10. Add 50 µL of Stop Solution to each well.
- 11. Read at 450 nm immediately.

Restrictions:


For Research Use only

Handling

Storage:	-20 °C
Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated
	freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is
	recommended to store at -80°C.

Expiry Date: 6 months

Images



ELISA

Image 1.

Human Serpin F2 concentration (ng/ml)