

Datasheet for ABIN2703466

Sclerostin ELISA Kit





Overview

Quantity:	96 tests
Target:	Sclerostin (SOST)
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	15-10000 pg/mL
Minimum Detection Limit:	15 pg/mL
Application:	ELISA

Product Details

Product Details	
Purpose:	Human SOST ELISA Kit for cell culture supernatants, plasma, and serum samples.
Sample Type:	Plasma, Cell Culture Supernatant, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA antibody pair detects human SOST. Other species not determined yet.
Sensitivity:	40 pg/mL
Characteristics:	 Strip plates and additional reagents allow for use in multiple experiments Quantitative protein detection Establishes normal range The best products for confirmation of antibody array data
Components:	Pre-Coated 96-well Strip Microplate

Product Details

- · Wash Buffer
- · Stop Solution
- · Assay Diluent(s)
- · Lyophilized Standard
- · Biotinylated Detection Antibody
- · Streptavidin-Conjugated HRP
- · TMB One-Step Substrate

Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2 μL to 1 μL volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 µL and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- · Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- · Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target:	Sclerostin (SOST)
Alternative Name:	SOST (SOST Products)
Background:	Gene Names: SOST UNQ2976/PR07455/PR07476
	Protein names: Sclerostin
Gene ID:	Protein names: Sclerostin 50964

Application Details

Application Details	
Application Notes:	Recommended Dilution for serum and plasma samples2 - 10 fold
Sample Volume:	100 μL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual.
	2. Add 100 µL of standard or sample to each well.
	3. Incubate 2.5 h at RT or O/N at 4 °C.
	4. Add 100 μL of prepared biotin antibody to each well.
	5. Incubate 1 h at RT.
	6. Add 100 μL of prepared Streptavidin solution to each well.
	7. Incubate 45 min at RT.

8. Add 100 μL of TMB One-Step Substrate Reagent to each well.

9. Incubate 30 min at RT.

10. Add 50 µL of Stop Solution to each well.

11. Read at 450 nm immediately.

Restrictions:

For Research Use only

6 months

Handling

Storage: -20 °C

Storage Comment: The entire kit may be stored at -20 °C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4 °C for up to 6 months. For extended storage, it is recommended to store at -80 °C.

Images

Expiry Date:

Assay Diluent Assay Diluent One of the second of the sec

ELISA

Image 1.