

Datasheet for ABIN2703492
Thyroperoxidase ELISA Kit[Go to Product page](#)

1 Image

Overview

Quantity:	96 tests
Target:	Thyroperoxidase (TPO)
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.3-100 ng/mL
Minimum Detection Limit:	0.3 ng/mL
Application:	ELISA

Product Details

Purpose:	Human Thyroid Peroxidase ELISA Kit for cell culture supernatants, EDTA treated plasma, and serum samples. Heparin and citrate are not recommended.
Sample Type:	Plasma, Cell Culture Supernatant, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA antibody pair detects human TPX. Other species not determined.
Sensitivity:	0.3 ng/mL
Characteristics:	<ul style="list-style-type: none">• Strip plates and additional reagents allow for use in multiple experiments• Quantitative protein detection• Establishes normal range• The best products for confirmation of antibody array data

Product Details

Components:	<ul style="list-style-type: none">• Pre-Coated 96-well Strip Microplate• Wash Buffer• Stop Solution• Assay Diluent(s)• Lyophilized Standard• Biotinylated Detection Antibody• Streptavidin-Conjugated HRP• TMB One-Step Substrate
-------------	--

Material not included:	<ul style="list-style-type: none">• Distilled or deionized water• Precision pipettes to deliver 2 µL to 1 µL volumes• Adjustable 1-25 µL pipettes for reagent preparation• 100 µL and 1 liter graduated cylinders• Tubes to prepare standard and sample dilutions• Absorbent paper• Microplate reader capable of measuring absorbance at 450nm• Log-log graph paper or computer and software for ELISA data analysis
------------------------	---

Target Details

Target:	Thyroperoxidase (TPO)
Alternative Name:	Thyroid Peroxidase (TPO Products)
Background:	Gene Names: TPO Protein names: Thyroid peroxidase (TPO) (EC 1.11.1.8)
Gene ID:	7173
UniProt:	P07202
Pathways:	Thyroid Hormone Synthesis

Application Details

Application Notes:	Recommended Dilution for serum and plasma samples2 fold
Sample Volume:	100 µL
Plate:	Pre-coated
Protocol:	<ol style="list-style-type: none">1. Prepare all reagents, samples and standards as instructed in the manual.2. Add 100 µL of standard or sample to each well.3. Incubate 2.5 h at RT or O/N at 4 °C.4. Add 100 µL of prepared biotin antibody to each well.

Application Details


- 5. Incubate 1 h at RT.
- 6. Add 100 µL of prepared Streptavidin solution to each well.
- 7. Incubate 45 min at RT.
- 8. Add 100 µL of TMB One-Step Substrate Reagent to each well.
- 9. Incubate 30 min at RT.
- 10. Add 50 µL of Stop Solution to each well.
- 11. Read at 450 nm immediately.

Sample Preparation:	Collect plasma using EDTA as an anticoagulant. Heparin and citrate are not recommended.
Restrictions:	For Research Use only

Handling

Storage:	-20 °C
Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C.
Expiry Date:	6 months

Images



ELISA

Image 1.