

Datasheet for ABIN2703505
TIMP2 ELISA Kit



[Go to Product page](#)

1 Image

Overview

Quantity:	96 tests
Target:	TIMP2
Reactivity:	Dog
Method Type:	Sandwich ELISA
Detection Range:	20-5000 pg/mL
Minimum Detection Limit:	20 pg/mL
Application:	ELISA

Product Details

Purpose:	Canine TIMP-2 ELISA Kit for Cell Culture Supernatants, Plasma, and Serum samples.
Sample Type:	Cell Culture Supernatant, Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	The antibody pair provided in this kit recognizes Canine Metalloproteinase inhibitor 2.
Sensitivity:	20 pg/mL
Characteristics:	• Strip plates and additional reagents allow for use in multiple experiments • Quantitative protein detection • Establishes normal range • The best products for confirmation of antibody array data
Components:	• Pre-Coated 96-well Strip Microplate

Product Details

- Wash Buffer
- Stop Solution
- Assay Diluent(s)
- Lyophilized Standard
- Biotinylated Detection Antibody
- Streptavidin-Conjugated HRP
- TMB One-Step Substrate

- Material not included:
- Distilled or deionized water
 - Precision pipettes to deliver 2 µL to 1 µL volumes
 - Adjustable 1-25 µL pipettes for reagent preparation
 - 100 µL and 1 liter graduated cylinders
 - Tubes to prepare standard and sample dilutions
 - Absorbent paper
 - Microplate reader capable of measuring absorbance at 450nm
 - Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target:	TIMP2
Alternative Name:	TIMP-2 (TIMP2 Products)
Background:	Gene Names: TIMP2 Protein names: Metalloproteinase inhibitor 2 (Tissue inhibitor of metalloproteinases 2) (TIMP-2)
Gene ID:	403633
UniProt:	Q9TTY1
Pathways:	cAMP Metabolic Process

Application Details

Application Notes:	Recommended Dilution for serum and plasma samples10 fold
Sample Volume:	100 µL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual. 2. Add 100 µL of standard or sample to each well. 3. Incubate 2.5 h at RT or O/N at 4 °C. 4. Add 100 µL of prepared biotin antibody to each well. 5. Incubate 1 h at RT.

Application Details

- 6. Add 100 µL of prepared Streptavidin solution to each well.
- 7. Incubate 45 min at RT.
- 8. Add 100 µL of TMB One-Step Substrate Reagent to each well.
- 9. Incubate 30 min at RT.
- 10. Add 50 µL of Stop Solution to each well.
- 11. Read at 450 nm immediately.

Restrictions: For Research Use only


Handling

Storage: -20 °C

Storage Comment: The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C.

Expiry Date: 6 months

Images



ELISA

Image 1.