.-online.com antibodies

Datasheet for ABIN2747901 SMAD2 ELISA Kit



Overview

Quantity:	96 tests
Target:	SMAD2
Binding Specificity:	pSer245, pSer250, pSer255, total
Reactivity:	Human, Rat, Mouse
Method Type:	Sandwich ELISA
Application:	ELISA

Product Details

Purpose:	Human/Mouse/Rat Phospho-SMAD2 (SER245/250/255) and Total SMAD2 Kit. This assay semi-quantitatively measures phophorylated SMAD2 (Ser245/250/255) and Total SMAD2 in lysate samples.
Sample Type:	Cell Lysate, Tissue Lysate
Analytical Method:	Semi-Quantitative
Detection Method:	Colorimetric
Specificity:	The antibody pair provided in this kit recognizes human, mouse, rat SMAD2 phosphorylated at site Serine-245/250/255 and Total SMAD2.
Characteristics:	 Simultaneously measure Phosphorylated protein and pan protein in one experiment (for normalization purpose) Screen numerous different cell lysates without performing a Western Blot analysis Minimal hands-on time, convenient, and non-radioactive material
Components:	Pre-Coated 96-well Strip Microplate

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN2747901 | 09/11/2023 | Copyright antibodies-online. All rights reserved.

	• Wash Buffer
	Anti-Phospho Antibody
	Anti-Pan Antibody
	HRP-Conjugated Secondary Antibody
	Streptavidin-Conjugated HRP
	Assay Diluent
	TMB One-Step Substrate
	Stop Solution
	Lysis Buffer
	Positive Control Sample
Material not included:	Distilled or deionized water
	 100 mL and 1 liter graduated cylinders
	Tubes to prepare sample dilutions
	Protease and Phosphatase inhibitors
	 Precision pipettes to deliver 2 µL to 1 mL volumes
	Adjustable 1-25 mL pipettes for reagent preparation
	Benchtop rocker or shaker
	 Microplate reader capable of measuring absorbance at 450 nm

Target Details

Target:	SMAD2
Alternative Name:	SMAD2 (SMAD2 Products)
Background:	Mothers against decapentaplegic homolog 2 (SMAD2) phosphorylated at Serine-245/250/255 and Total SMAD2
Gene ID:	4087
UniProt:	Q15796
Pathways:	Cell Division Cycle, Hormone Transport, Chromatin Binding, Protein targeting to Nucleus

Application Details

Sample Volume:	100 µL
Plate:	Pre-coated
Protocol:	 Prepare all reagents and samples as instructed in the manual. Add 100 μL of sample or positive control to each well. Incubate 2.5 h at RT or O/N at 4 °C. Add 100 μL of prepared primary antibody to each well.

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/3 | Product datasheet for ABIN2747901 | 09/11/2023 | Copyright antibodies-online. All rights reserved.

	 5. Incubate 1 h at RT. 6. Add 100 μL of prepared 1X HRP-Streptavidin to each well. 7. Incubate 1 h at RT. 8. Add 100 μL of TMB One-Step Substrate Reagent to each well. 9. Incubate 30 min at RT. 10. Add 50 μL of Stop Solution to each well. 11. Read at 450 nm immediately.
Assay Procedure:	Prepare all reagents and samples as instructed in the manual.
	Add 100 μ L of sample or positive control to each well.
	Incubate 2.5 h at RT or O/N at 4 °C.
	Add 100 μ L of prepared primary antibody to each well.
	Incubate 1 h at RT.
	Add 100 μ L of prepared 1X HRP-Streptavidin to each well.
	Incubate 1 h at RT.
	Add 100 μ L of TMB One-Step Substrate Reagent to each well.
	Incubate 30 min at RT.
	Add 50 μ L of Stop Solution to each well.
	Read at 450 nm immediately.
Restrictions:	For Research Use only

Handling

Storage:	-20 °C
Storage Comment:	Upon receipt, the kit should be stored at -20 $^\circ$ C. Please use within 6 months from the date of
	shipment. After initial use, Wash Buffer Concentrate (Item B), Assay Diluent (Item E), TMB One-
	Step Substrate Reagent (Item H), HRP-Streptavidin (Item G), Stop Solution (Item I) and Cell
	Lysate Buffer (Item J) should be stored at 4 °C to avoid repeated freeze-thaw cycles. Return
	unused wells to the pouch containing desiccant pack, reseal along entire edge and store at -20
	°C. Reconstituted Positive Control (Item K) should be stored at -70 °C.

Expiry Date:

6 months