antibodies .- online.com





Datasheet for ABIN2748385

MSK2 ELISA Kit



()	1/0	r\ /1	014	
()	ve	I V I	-v	V

Quantity:	96 tests	
Target:	MSK2 (RPS6KA4)	
Binding Specificity:	pSer360	
Reactivity:	Human	
Method Type:	Sandwich ELISA	
Application:	ELISA	
Product Details		
Purpose:	Human Phospho-MSK2 (S360) ELISA Kit. This assay semi-quantitatively measures	
	phophorylated MSK2 (Ser360) in lysate samples.	
Sample Type:	Cell Lysate, Tissue Lysate	
Analytical Method:	Semi-Quantitative	
Detection Method:	Colorimetric	
Specificity:	The antibody pair provided in this kit recognizes human MSK2 phosphorylated at site Serine- 360.	
Characteristics:	Rapidly measure phosphorylated protein in lysates	
	 Screen numerous different cell lysates without performing a Western Blot analysis Minimal hands-on time, convenient, and non-radioactive material 	
Components:	Pre-Coated 96-well Strip Microplate	
	Wash BufferAnti-Phospho Antibody	

Product Details

- · HRP-Conjugated Secondary Antibody
- · Assay Diluent
- · TMB One-Step Substrate
- · Stop Solution
- · Lysis Buffer
- Positive Control Sample

Material not included:

- · Distilled or deionized water
- 100 mL and 1 liter graduated cylinders
- Tubes to prepare sample dilutions
- · Protease and Phosphatase inhibitors
- Precision pipettes to deliver 2 µL to 1 mL volumes
- Adjustable 1-25 mL pipettes for reagent preparation
- · Benchtop rocker or shaker
- Microplate reader capable of measuring absorbance at 450 nm

Target Details

Target:	MSK2 (RPS6KA4)	
Alternative Name:	MSK2 (RPS6KA4 Products)	
Background:	Mitogen- and stress-activated protein kinase 2 (MSK2) phosphorylated at Serine-360	
Gene ID:	8986	
UniProt:	075676	

Application Details

Sample Volume:	100 μL	
Plate:	Pre-coated	
Protocol:	1. Prepare all reagents and samples as instructed in the manual.	
	2. Add 100 µL of sample or positive control to each well.	
	3. Incubate 2.5 h at RT or O/N at 4 °C.	
	4. Add 100 μL of prepared primary antibody to each well.	
	5. Incubate 1 h at RT.	
	6. Add 100 μL of prepared 1X HRP-Streptavidin to each well.	
	7. Incubate 1 h at RT.	
	8. Add 100 µL of TMB One-Step Substrate Reagent to each well.	
	9. Incubate 30 min at RT.	
	10. Add 50 μL of Stop Solution to each well.	
	11. Read at 450 nm immediately.	

Application Details

Expiry Date:

6 months

Assay Procedure:	Prepare all reagents and samples as instructed in the manual.	
	Add 100 μ L of sample or positive control to each well.	
	Incubate 2.5 h at RT or O/N at 4 °C.	
	Add 100 μL of prepared primary antibody to each well.	
	Incubate 1 h at RT.	
	Add 100 μL of prepared 1X HRP-Streptavidin to each well.	
	Incubate 1 h at RT.	
	Add 100 μL of TMB One-Step Substrate Reagent to each well.	
	Incubate 30 min at RT.	
	Add 50 μL of Stop Solution to each well.	
	Read at 450 nm immediately.	
Restrictions:	For Research Use only	
Handling		
Storage:	-20 °C	
Storage Comment:	Upon receipt, the kit should be stored at -20 °C. Please use within 6 months from the date of	
ū	shipment. After initial use, Wash Buffer Concentrate (Item B), Assay Diluent (Item E), TMB One-	
	Step Substrate Reagent (Item H), HRP-Streptavidin (Item G), Stop Solution (Item I) and Cell	
	Lysate Buffer (Item J) should be stored at 4 °C to avoid repeated freeze-thaw cycles. Return	
	unused wells to the pouch containing desiccant pack, reseal along entire edge and store at -20	
	°C. Reconstituted Positive Control (Item K) should be stored at -70 °C.	