## .-online.com antibodies

# Datasheet for ABIN2748476 CDKN1B ELISA Kit



#### Overview

Quantity:	96 tests
Target:	CDKN1B
Reactivity:	Rat
Method Type:	Sandwich ELISA
Application:	ELISA

#### Product Details

Purpose:	Custom Rat p27/Kip1 (CDKN1B) ELISA Kit.
Sample Type:	Cell Culture Supernatant, Cell Lysate, Plasma, Serum, Tissue Lysate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	The antibody pair provided in this kit recognizes Rat p27/Kip1 (CDKN1B)
Characteristics:	<ul> <li>Strip plates and additional reagents allow for use in multiple experiments</li> <li>Quantitative protein detection</li> <li>Establishes normal range</li> <li>The best products for confirmation of antibody array data</li> </ul>
Components:	<ul> <li>Pre-Coated 96-well Strip Microplate</li> <li>Wash Buffer</li> <li>Stop Solution</li> <li>Assay Diluent(s)</li> <li>Lyophilized Standard</li> <li>Biotinylated Detection Antibody</li> </ul>

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN2748476 | 09/11/2023 | Copyright antibodies-online. All rights reserved.

	<ul><li>Streptavidin-Conjugated HRP</li><li>TMB One-Step Substrate</li></ul>
Material not included:	Distilled or deionized water
	<ul> <li>Precision pipettes to deliver 2 µL to 1 µL volumes</li> </ul>
	<ul> <li>Adjustable 1-25 µL pipettes for reagent preparation</li> </ul>
	<ul> <li>100 µL and 1 liter graduated cylinders</li> </ul>
	Tubes to prepare standard and sample dilutions
	Absorbent paper
	Microplate reader capable of measuring absorbance at 450nm
	Log-log graph paper or computer and software for ELISA data analysis

### Target Details

Target:CDKN1BAlternative Name:p27/Kip1 (CDKN1B Products)Background:P27/Kip1 (CDKN1B)Gene ID:83571UniProt:008769Pathways:Cell Division Cycle, Fc-epsilon Receptor Signaling Pathway, EGFR Signaling Pathway, Neurotrophin Signaling Pathway, Positive Regulation of Peptide Hormone Secretion, Negative		
Background:P27/Kip1 (CDKN1B)Gene ID:83571UniProt:008769Pathways:Cell Division Cycle, Fc-epsilon Receptor Signaling Pathway, EGFR Signaling Pathway,	Target:	CDKN1B
Gene ID:83571UniProt:008769Pathways:Cell Division Cycle, Fc-epsilon Receptor Signaling Pathway, EGFR Signaling Pathway,	Alternative Name:	p27/Kip1 (CDKN1B Products)
UniProt:     008769       Pathways:     Cell Division Cycle, Fc-epsilon Receptor Signaling Pathway, EGFR Signaling Pathway,	Background:	P27/Kip1 (CDKN1B)
Pathways:       Cell Division Cycle, Fc-epsilon Receptor Signaling Pathway, EGFR Signaling Pathway,	Gene ID:	83571
	UniProt:	008769
Neurotrophin Signaling Pathway, Positive Regulation of Peptide Hormone Secretion, Negati	Pathways:	Cell Division Cycle, Fc-epsilon Receptor Signaling Pathway, EGFR Signaling Pathway,
		Neurotrophin Signaling Pathway, Positive Regulation of Peptide Hormone Secretion, Negative
Regulation of Hormone Secretion, Sensory Perception of Sound, Mitotic G1-G1/S Phases, D		Regulation of Hormone Secretion, Sensory Perception of Sound, Mitotic G1-G1/S Phases, DNA
Replication, Positive Regulation of Endopeptidase Activity, Synthesis of DNA, Autophagy		Replication, Positive Regulation of Endopeptidase Activity, Synthesis of DNA, Autophagy

#### Application Details

Sample Volume:	100 µL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual.
	2. Add 100 $\mu$ L of standard or sample to each well.
	3. Incubate 2.5 h at RT or O/N at 4 °C.
	4. Add 100 $\mu$ L of prepared biotin antibody to each well.
	5. Incubate 1 h at RT.
	6. Add 100 $\mu$ L of prepared Streptavidin solution to each well.
	7. Incubate 45 min at RT.
	8. Add 100 μL of TMB One-Step Substrate Reagent to each well.
	9. Incubate 30 min at RT.

Application Details	
	10. Add 50 μL of Stop Solution to each well. 11. Read at 450 nm immediately.
Assay Procedure:	Prepare all reagents, samples and standards as instructed in the manual.Add 100 $\mu$ L of standard or sample to each well.Incubate 2.5 h at RT or O/N at 4 °C.Add 100 $\mu$ L of prepared biotin antibody to each well.Incubate 1 h at RT.Add 100 $\mu$ L of prepared Streptavidin solution to each well.Incubate 45 min at RT.Add 100 $\mu$ L of TMB One-Step Substrate Reagent to each well.Incubate 30 min at RT.Add 50 $\mu$ L of Stop Solution to each well.Read at 450 nm immediately.
Restrictions:	For Research Use only
Handling	
Storage:	-20 °C
Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C.
Expiry Date:	6 months