

## Datasheet for ABIN2748589

# **CCL5 ELISA Kit**



### Overview

Quantity:	96 tests
Target:	CCL5
Reactivity:	Pig
Method Type:	Sandwich ELISA
Detection Range:	46.9-3000 pg/mL
Minimum Detection Limit:	46.9 pg/mL
Application:	ELISA

#### Product Details

Product Details	
Purpose:	Porcine RANTES (CCL5) ELISA Kit for serum, plasma, and cell culture supernatant samples.
Sample Type:	Plasma, Cell Culture Supernatant, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	Cross Reactivity: This ELISA kit shows no cross-reactivity with the following cytokines tested: porcine Angiopoietin-1, IL-17F, MIF, Osteoprotegerin (TNFRSF11B), PDGF-BB, TGF alpha, TIMP-1, TIMP-2, VEGF-A
Characteristics:	<ul> <li>Strip plates and additional reagents allow for use in multiple experiments</li> <li>Quantitative protein detection</li> <li>Establishes normal range</li> <li>The best products for confirmation of antibody array data</li> </ul>

### **Product Details**

#### Components:

- · Pre-Coated 96-well Strip Microplate
- · Wash Buffer
- · Stop Solution
- · Assay Diluent(s)
- · Lyophilized Standard
- · Biotinylated Detection Antibody
- · Streptavidin-Conjugated HRP
- TMB One-Step Substrate

### Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2 μL to 1 μL volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 µL and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- · Absorbent paper
- · Microplate reader capable of measuring absorbance at 450nm
- · Log-log graph paper or computer and software for ELISA data analysis

## **Target Details**

Target:	CCL5
Alternative Name:	RANTES (CCL5 Products)
Background:	RANTES (CCL5)
UniProt:	Q29288
Pathways:	Cellular Response to Molecule of Bacterial Origin, Regulation of G-Protein Coupled Receptor Protein Signaling, Smooth Muscle Cell Migration

### **Application Details**

Application Notes:	Recommended Dilution for serum and plasma samples2 fold
Sample Volume:	100 μL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual.
	2. Add 100 µL of standard or sample to each well.
	3. Incubate 2.5 h at RT or O/N at 4 °C.
	4. Add 100 μL of prepared biotin antibody to each well.
	5. Incubate 1 h at RT.
	6. Add 100 $\mu$ L of prepared Streptavidin solution to each well.

### **Application Details**

Expiry Date:

6 months

7. Incubate 45 min at RT. 8. Add 100 µL of TMB One-Step Substrate Reagent to each well. 9. Incubate 30 min at RT. 10. Add 50 µL of Stop Solution to each well. 11. Read at 450 nm immediately. Assay Procedure: Prepare all reagents, samples and standards as instructed in the manual. Add 100 µL of standard or sample to each well. Incubate 2.5 h at RT or O/N at 4 °C. Add 100 µL of prepared biotin antibody to each well. Incubate 1 h at RT.Add 100 µL of prepared Streptavidin solution to each well.Incubate 45 min at RT.Add 100  $\mu L$  of TMB One-Step Substrate Reagent to each well.Incubate 30 min at RT.Add 50 μL of Stop Solution to each well.Read at 450 nm immediately. Restrictions: For Research Use only Handling -20 °C Storage: Storage Comment: The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C.