

[Go to Product page](#)

Datasheet for ABIN2871394 Endomorphin 1 FLISA Kit

Overview

Quantity:	96 tests
Target:	Endomorphin 1 (EM1)
Reactivity:	Human
Method Type:	Competition ELISA
Detection Range:	29.7-940 pg/mL
Minimum Detection Limit:	29.7 pg/mL
Application:	ELISA

Product Details

Analytical Method:	Quantitative
Detection Method:	Fluorometric

Target Details

Target:	Endomorphin 1 (EM1)
Alternative Name:	Endomorphin-1 (EM1 Products)

Application Details

Application Notes:	Optimal working dilution should be determined by the investigator.
Plate:	Pre-coated
Protocol:	The immunoplate in this kit is pre-coated with a secondary antibody and the nonspecific

Application Details

binding sites are blocked. The secondary antibody can bind to the Fc fragment of the primary antibody (peptide antibody) whose Fab fragment will be competitively bound by both biotinylated peptide and peptide standard or targeted peptide in samples. The biotinylated peptide interacts with streptavidin-horseradish peroxidase (SA-HRP) which catalyzes the substrate. The fluorescence intensity is directly proportional to the amount of biotinylated peptide-SA-HRP complex but inversely proportional to the amount of the peptide in standard solutions or samples. This is due to the competitive binding of the biotinylated peptide with the standard peptide or samples to the peptide antibody (primary antibody). A standard curve of known concentration can be established accordingly. The unknown concentration in samples can be determined by extrapolation to this standard curve.

Restrictions: For Research Use only

Handling

Storage: 4 °C