

Datasheet for ABIN2943983

CECR1 CLIA Kit



[Go to Product page](#)

Overview

Quantity:	96 tests
Target:	CECR1
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.312 ng/mL - 20 ng/mL
Minimum Detection Limit:	0.312 ng/mL
Application:	ELISA

Product Details

Purpose:	Human Cat Eye Syndrome Chromosome Region, Candidate 1 (CECR1) CLIA Kit is a Sandwich CLIA Kit for use with Serum, plasma and other biological fluids.
Sample Type:	Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Chemiluminescent
Sensitivity:	< 0.118 ng/mL

Target Details

Target:	CECR1
Alternative Name:	Cat Eye Syndrome Chromosome Region, Candidate 1 (CECR1 Products)
Pathways:	Ribonucleoside Biosynthetic Process

Application Details

Application Notes: Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout. Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user.

Standard Form: Lyophilized

Plate: Pre-coated

Restrictions: For Research Use only

Handling

Storage: 4 °C/-20 °C

Storage Comment: Upon receipt, store the kit according to the storage instruction in the kit's manual.

Expiry Date: 6 months