

Datasheet for ABIN2944054  
**D Amino Acid Oxidase CLIA Kit**



[Go to Product page](#)

## Overview

Quantity:	96 tests
Target:	D Amino Acid Oxidase (DAO)
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	1.56 ng/mL - 100 ng/mL
Minimum Detection Limit:	1.56 ng/mL
Application:	ELISA

## Product Details

Purpose:	Human D-Amino Acid Oxidase CLIA Kit is a Sandwich CLIA Kit for use with Serum, plasma, tissue homogenates and other biological fluids.
Sample Type:	Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Chemiluminescent
Sensitivity:	< 0.56 ng/mL

## Target Details

Target:	D Amino Acid Oxidase (DAO)
Alternative Name:	D-Amino Acid Oxidase ( <a href="#">DAO Products</a> )

## Application Details

---

**Application Notes:**                      **Stability:** The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout.

**Recommended dilutions:** Optimal dilutions/concentrations should be determined by the end user.

**Standard Form:** Lyophilized

---

**Plate:**                                        Pre-coated

---

**Restrictions:**                              For Research Use only

---

## Handling

---

**Storage:**                                    4 °C/-20 °C

---

**Storage Comment:**                      Upon receipt, store the kit according to the storage instruction in the kit's manual.

---

**Expiry Date:**                              6 months

---