

Datasheet for ABIN2944398 **LEFTY1 CLIA Kit**



[Go to Product page](#)

Overview

Quantity: 96 tests

Target: LEFTY1

Reactivity: Human

Method Type: Sandwich ELISA

Detection Range: 0.781 ng/mL - 50 ng/mL

Minimum Detection Limit: 0.781 ng/mL

Application: ELISA

Product Details

Purpose: Human Left/Right Determination Factor 1 (LEFTY1) CLIA Kit is a Sandwich CLIA Kit for use with Tissue homogenates, cell lysates, cell culture supernates and other biological fluids.

Sample Type: Cell Culture Supernatant, Cell Lysate, Tissue Homogenate

Analytical Method: Quantitative

Detection Method: Chemiluminescent

Sensitivity: < 0.31 ng/mL

Target Details

Target: LEFTY1

Alternative Name: Left/Right Determination Factor 1 ([LEFTY1 Products](#))

Application Details

Application Notes: **Stability:** The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout. **Recommended dilutions:** Optimal dilutions/concentrations should be determined by the end user.
Standard Form: Lyophilized

Plate: Pre-coated

Restrictions: For Research Use only

Handling

Storage: 4 °C/-20 °C

Storage Comment: Upon receipt, store the kit according to the storage instruction in the kit's manual.

Expiry Date: 6 months
