

[Go to Product page](#)

Datasheet for ABIN2950915

## Prothrombin Fragment 1+2 ELISA Kit

### Overview

|                          |                                 |
|--------------------------|---------------------------------|
| Quantity:                | 96 tests                        |
| Target:                  | Prothrombin Fragment 1+2 (F1+2) |
| Reactivity:              | Human                           |
| Method Type:             | Sandwich ELISA                  |
| Detection Range:         | 46.875 pg/mL - 3000 pg/mL       |
| Minimum Detection Limit: | 46.875 pg/mL                    |
| Application:             | ELISA                           |

### Product Details

|                    |  |
|--------------------|--|
| Purpose:           | Human Prothrombin Fragment 1+2 ELISA Kit is an ELISA Kit against Prothrombin Fragment 1+2. |
| Sample Type:       | Plasma, Serum  |
| Analytical Method: | Quantitative   |
| Detection Method:  | Colorimetric   |
| Sensitivity:       | 28.125 pg/mL   |

### Target Details

|                   |  |
|-------------------|--|
| Target:           | Prothrombin Fragment 1+2 (F1+2)                            |
| Alternative Name: | Prothrombin Fragment 1+2 ( <a href="#">F1+2 Products</a> ) |

## Application Details

---

**Application Notes:**                      **Stability:** The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout. **Recommended dilutions:** Optimal dilutions/concentrations should be determined by the end user.  
**Standard Form:** Lyophilized

---

**Plate:**    Pre-coated

---

**Restrictions:**                                      For Research Use only

---

## Handling

---

**Storage:**    4 °C/-20 °C

---

**Storage Comment:**                                      Upon receipt, store the kit according to the storage instruction in the kit's manual.

---

**Expiry Date:**    6 months

---