

Datasheet for ABIN2952042

TREM1 ELISA Kit



| _ | verview | | | | | |
|---|---------|-------------|----|----|-------------|-----|
| | 1// | \triangle | r۱ | /1 | \triangle | Λ/ |
| | ' V ' | | ΙV | | | v v |

| Quantity: | 96 tests |
|---|--|
| Target: | TREM1 |
| Reactivity: | Human |
| Method Type: | Sandwich ELISA |
| Detection Range: | 31.25 pg/mL - 2000 pg/mL |
| Minimum Detection Limit: | 31.25 pg/mL |
| Application: | ELISA |
| Product Details | |
| Purpose: | Human Soluble Triggering Receptor Expressed On Myeloid Cells 1 (TREM1) ELISA Kit is an |
| | FUICA Kit a gainst Caluble Triggering December Fugures and On Myselaid Calls 1 |
| | ELISA Kit against Soluble Triggering Receptor Expressed On Myeloid Cells 1. |
| Sample Type: | Plasma, Serum |
| Sample Type: Analytical Method: | |
| | Plasma, Serum |
| Analytical Method: | Plasma, Serum Quantitative |
| Analytical Method: Detection Method: | Plasma, Serum Quantitative Colorimetric |
| Analytical Method: Detection Method: Sensitivity: | Plasma, Serum Quantitative Colorimetric |

Application Details

| Application Notes: | Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less | | | |
|--------------------|---|--|--|--|
| | than 5 % within the expiration date under appropriate storage conditions. To minimize | | | |
| | performance fluctuations, operation procedures and lab conditions should be strictly controlled. | | | |
| | It is also strongly suggested that the whole assay is performed by the same user throughout. | | | |
| | Recommended dilutions: Optimal dilutions/concentrations should be determined by the end | | | |
| | user. | | | |
| Plate: | Pre-coated | | | |
| Restrictions: | For Research Use only | | | |
| | | | | |
| Handling | | | | |
| Storage: | 4 °C/-20 °C | | | |
| Storage Comment: | Upon receipt, store the kit according to the storage instruction in the kit's manual. | | | |