

Datasheet for ABIN2952503

ERCC2 ELISA Kit



| _ | | | | | |
|---|---|---|----|----|---|
| | W | 0 | rv | 10 | W |

| Overview | | |
|--------------------------|---|--|
| Quantity: | 96 tests | |
| Target: | ERCC2 | |
| Reactivity: | Human | |
| Method Type: | Sandwich ELISA | |
| Detection Range: | 0.156 ng/mL - 10 ng/mL | |
| Minimum Detection Limit: | 0.156 ng/mL | |
| Application: | ELISA | |
| Product Details | | |
| Purpose: | Human Xeroderma Pigmentosum, Complementation Group D ELISA Kit is a ELISA Kit against | |
| | Xeroderma Pigmentosum, Complementation Group D. | |
| Sample Type: | Cell Lysate, Tissue Homogenate | |
| Analytical Method: | Quantitative | |
| Detection Method: | Colorimetric | |
| Sensitivity: | < 0.057 ng/mL | |
| Characteristics: | Xeroderma Pigmentosum, Complementation Group D | |
| Target Details | | |
| Target: | ERCC2 | |
| Alternative Name: | Xeroderma Pigmentosum, Complementation Group D (ERCC2 Products) | |
| | | |

Target Details

| Pathways: | DNA Damage Repair | | |
|---------------------|---|--|--|
| Application Details | | | |
| Application Notes: | Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less | | |
| | than 5 % within the expiration date under appropriate storage conditions. To minimize | | |
| | performance fluctuations, operation procedures and lab conditions should be strictly controlled. | | |
| | It is also strongly suggested that the whole assay is performed by the same user throughout. | | |
| | Recommended dilutions: Optimal dilutions/concentrations should be determined by the end | | |
| | user. | | |
| | Standard Form: Lyophilized | | |
| Plate: | Pre-coated | | |
| Restrictions: | For Research Use only | | |
| Handling | | | |
| Storage: | 4 °C/-20 °C | | |
| Storage Comment: | Upon receipt, store the kit according to the storage instruction in the kit's manual. | | |
| Expiry Date: | 6 months | | |