

Datasheet for ABIN364986

ADIPOQ ELISA Kit[Go to Product page](#)**1** Validation**2** Images**11** Publications

Overview

Quantity: 96 tests

Target: ADIPOQ

Reactivity: Human

Method Type: Sandwich ELISA

Detection Range: 1.562-100 ng/mL

Minimum Detection Limit: 1.562 ng/mL

Application: ELISA

Product Details

Purpose: For the quantitative determination of human adiponectin (ADP) concentrations in serum, plasma, cell culture supernates, tissue homogenates, urine.

Sample Type: Serum, Plasma, Cell Culture Supernatant, Tissue Homogenate, Urine

Analytical Method: Quantitative

Detection Method: Colorimetric

Specificity: This assay has high sensitivity and excellent specificity for detection of human ADP.

Cross-Reactivity (Details): Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between the target antigen and all analogues for other species. Therefore, cross reaction may still exist.

Sensitivity: 1.102 ng/mL

Components:

- Assay plate (12 × 8 coated Microwells)

Product Details

- Standard (freeze dried)
- Biotin-antibody (100 × concentrate)
- HRP-avidin (100 × concentrate)
- Biotin-antibody Diluent
- HRP-avidin Diluent
- Sample Diluent
- Wash Buffer (25 × concentrate)
- TMB Substrate
- Stop Solution
- Adhesive Strip (for 96 wells)
- Instruction manual

Target Details

Target:	ADIPOQ
Alternative Name:	Adiponectin (ADP) (ADIPOQ Products)
Background:	Synonyms: ACDC, ACRP30, ADIPQTL1, ADPN, APM-1, APM1, GBP28, adiponectin, OTTHUMP00000210047 OTTHUMP00000210048 adipocyte, C1Q and collagen domain containing adipocyte, C1Q and collagen domain-containing adipo
HGNC:	13633
UniProt:	Q15848

Application Details

Application Notes:	<ul style="list-style-type: none">• The supplier is only responsible for the kit itself, but not for the samples consumed during the assay. The user should calculate the possible amount of the samples used in the whole test. Please reserve sufficient samples in advance.• Samples to be used within 5 days may be stored at 2-8°C, otherwise samples must be stored at -20°C (≤ 1 month) or -80°C (≤ 2 months) to avoid loss of bioactivity and contamination.• Grossly hemolyzed samples are not suitable for use in this assay.• If the samples are not indicated in the manual, a preliminary experiment to determine the validity of the kit is necessary.• Please predict the concentration before assaying. If values for these are not within the range of the standard curve, users must determine the optimal sample dilutions for their particular experiments.• Tissue or cell extraction samples prepared by chemical lysis buffer may cause unexpected ELISA results due to the impacts of certain chemicals.• Owing to the possibility of mismatching between antigens from another resource and antibodies used in this supplier's kits (e.g., antibody targets conformational epitope rather than linear epitope), some native or recombinant proteins from other manufacturers may not
--------------------	---

be recognized by this supplier's products.

- Influenced by factors including cell viability, cell number and cell sampling time, samples from cell culture supernatant may not be recognized by the kit.
- Fresh samples without long time storage are recommended for the test. Otherwise, protein degradation and denaturalization may occur in those samples and finally lead to wrong results.

Comment:

Detection wavelength: 450 nm

Information on standard material:

Depending on the antigen to be detected, standards can be either native or recombinant protein. The recombinant proteins are being expressed in CHO cells in most cases. Please inquire for more information. The formulation of auxiliary material in the standard is considered proprietary information, however it does not contain any poisonous substance. Proclin 300 (1:3000) is used as preservative.

Information on reagents:

In most cases the stop solution provided is 1 N H₂SO₄. The formulation of wash solution is proprietary information. None of the components contain (sodium) azide, thimerosal, 2-mercaptoethanol (2-ME) or any other poisonous materials. For the sandwich method kits, the sample diluent, antibody diluent, enzyme diluent and standard all contain BSA.

Information on antibodies:

The antibodies provided in different kits vary in regards to clonality and host. Some antibodies are affinity purified, some are Protein A

Sample Volume:

100 µL

Assay Time:

1 - 4.5 h

Plate:

Pre-coated

Protocol:

This assay employs the quantitative sandwich enzyme immunoassay technique. Antibody specific for ADP has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any ADP present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for ADP is added to the wells. After washing, avidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of ADP bound in the initial step. The color development is stopped and the intensity of the color is measured.

Application Details

Reagent Preparation:

- **Biotin-antibody (1×)** - Centrifuge the vial before opening.
Biotin-antibody requires a 100-fold dilution. The suggested dilution is 10µL of Biotin-antibody + 990µL of Biotin-antibody Diluent.
- **HRP-avidin (1×)** - Centrifuge the vial before opening.
HRP-avidin requires a 100-fold dilution. The suggested dilution is 10µL of HRP-avidin + 990µL of HRP-avidin Diluent.
- **Wash Buffer (1×)** - If crystals have formed in the concentrate, warm up to room temperature and mix gently until the crystals have completely dissolved. Dilute 20mL of Wash Buffer Concentrate (25×) into deionized or distilled water to prepare 500mL of Wash Buffer (1×).
- **Standard** - Centrifuge the standard vial at 6000-10000rpm for 30s.
Reconstitute the Standard with 1ml of Sample Diluent. Do not substitute other diluents. This reconstitution produces a stock solution of 200pg/mL. Mix the standard to ensure complete reconstitution and allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions.
Pipette 250µL of Sample Diluent into each tube. Use the stock solution to produce a 2-fold dilution series. Mix each tube thoroughly before the next transfer. The undiluted Standard serves as the high standard (200pg/mL). Sample Diluent serves as the zero standard (0ng/mL).

Note:

- Kindly use graduated containers to prepare the reagent. Please don't prepare the reagent directly in the Diluent vials provided in the kit.
- Bring all reagents to room temperature (18-25°C) before use for 30 min.
- Prepare fresh standard for each assay. Use within 4 hours and discard after use.
- Making serial dilution in the wells directly is not permitted.
- Please carefully reconstitute Standards according to the instruction. Avoid foaming and mix gently until the crystals have completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL when pipetting.
- It is recommended to use distilled water to prepare reagents and samples. Using contaminated water or container for reagent preparation will influence detection result.

Assay Precision:

Intra-assay precision (precision within an assay): Three samples of known concentration were tested twenty times on one plate to assess precision.

Inter-assay precision (precision between assays): Three samples of known concentration were tested in twenty assays to assess precision.

- Intra-assay: CV% less than 8%
- Inter-assay: CV% less than 10%

Restrictions:

For Research Use only

Handling

Precaution of Use:	The Stop Solution provided with this kit is an acid solution. Wear eye, hand, face and clothing protection when using this material.
Handling Advice:	<ul style="list-style-type: none">• The kit should not be used beyond the expiration date on the kit label.• Do not mix or substitute reagents with those from other lots or sources.• If samples generate values higher than the highest standard, dilute the samples with Sample Diluent and repeat the assay.• Any variation in Sample Diluent, operator, pipetting technique, washing technique, incubation time/temperature and kit age can cause variation in binding.• This assay is designed to eliminate interference by soluble receptors, binding proteins and other factors present in biological samples. Until all factors have been tested in the Immunoassay, the possibility of interference cannot be excluded.
Storage:	4 °C/-20 °C
Storage Comment:	For unopened kit: All the reagents should be kept according to the labels on vials.
Expiry Date:	6 months

Publications

Product cited in:	<p>Li, Zhan, Liao, Wang, Liu: "Plasma levels of adiponectin and chemerin are associated with early stage of atherosclerosis in older adults with type 2 diabetes mellitus." in: Aging medicine (Milton (N.S.W)), Vol. 2, Issue 4, pp. 198-206, (2019) (PubMed).</p> <p>Li, Wang, Yang, Yao, Zhang, Xiao, Zhang, Wang, Xu, Wang: "Interaction between β-hexachlorocyclohexane and ADIPOQ genotypes contributes to the risk of type 2 diabetes mellitus in East Chinese adults." in: Scientific reports, Vol. 6, pp. 37769, (2018) (PubMed).</p> <p>Chen, Huang, Lin, Chen, Zhao: "Relationship between serum adiponectin and bone mineral density in male patients with obstructive sleep apnea syndrome." in: Sleep & breathing = Schlaf & Atmung, Vol. 21, Issue 2, pp. 557-564, (2018) (PubMed).</p> <p>Torabi, Farahani, Safakish, Ramezankhani, Dehghan: "Evaluation of motor proficiency and adiponectin in adolescent students with attention deficit hyperactivity disorder after high-intensity intermittent training." in: Psychiatry research, Vol. 261, pp. 40-44, (2018) (PubMed).</p> <p>Liu, Zhong, Huang: "Elevated Plasma miR-29a Levels Are Associated with Increased Carotid Intima-Media Thickness in Atherosclerosis Patients." in: The Tohoku journal of experimental medicine, Vol. 241, Issue 3, pp. 183-188, (2017) (PubMed).</p>
-------------------	---

There are more publications referencing this product on: [Product page](#)

Validation report #029594 for ELISA (ELISA)





Image 1.



ELISA

Image 2. Typical standard curve



Successfully validated (ELISA (ELISA))

by [Alamo Laboratories Inc](#)

Report Number: 029593

Date: Feb 09 2014

Lot Number: T30095901

Method validated: ELISA (ELISA)

Positive Control: [Human serum](#)

Negative Control: Goat serum

Notes: Signal was detected in positive control serum but not in negative control serum.

Primary Antibody: - Antigen: Human Adiponectin (ADP) - Catalog Number: ABIN364986 - Supplier: Cusabio -
Supplier Catalog number: CSB-E07270h - Lot Number: T30095901

Controls:

- Positive control: Serum from normal adult human (specimen known to contain the target protein).
- Negative control: Serum from normal goat (specimens known to not contain the target protein).
- Standard curve: Serial two-fold dilutions from 100 ng/ml [100, 50, 25, 12.5, 6.25, 3.13, 1.56, 0] were generated from the standard provided in the kit using standard/sample diluent buffer.
- Spike control: Standard diluted in standard/PBS diluent buffer [6.25 and 0].

Protocol:


- All reagents in the ELISA kit were brought up to room temperature (RT) before use.
- 100 µL of standard or sample were added to wells in ELISA plate pre-coated with capture antibody. All samples and standards were assayed in triplicate.
- The plate was covered with sealer (provided in kit) and incubated for 2 hours at 37°C. Unbound material was aspirated but the wells were NOT Washed.
- 100 µL of Biotin-Antibody (diluted 1:100 in "Biotin-Antibody Diluent") was added to each well. Plate was covered with sealer (provided in kit) and incubated for 1 hour at 37°C. Unbound Biotin-Antibody was removed from each well and plate was washed three times with 350 µL of wash buffer (provided in the kit). After the last wash the plate was inverted against clean absorbent paper to remove any remaining liquid.
- 100 µL of HRP-Avidin Conjugate (diluted 1:100 in "HRP-Avidin Diluent") was added to each well. Plate was covered with sealer (provided in kit) and incubated for 1 hour at 37°C.
- Unbound HRP-Avidin was removed by washing five times with 350 µL of wash buffer (provided in the kit). After the last wash the plate was inverted and blotted against clean absorbent paper to remove any remaining liquid.
- 90 µL of TMB substrate was added to wells and the plate was covered with a new plate

sealer. The plate was gently tapped to ensure mixing and incubated for 30 min at 37°C in the dark.

- After 30 min, when an apparent gradient appeared in the standard wells, the reaction was terminated by adding 50 µL of Stop Solution to each well.
- The optical density (OD value) of each well was read using a microplate reader set to 450 nm.
- The triplicate readings for each sample were averaged and the average zero standard optical density subtracted to yield 'corrected absorbance at 450 nm'. A standard curve was generated by plotting the mean OD value for each standard on the X-axis against the concentration on the Y-axis using Excel. Standard curve was generated by regression analysis with four-parameter logistic.
- An equation ($y = 4.2198x^4 - 16.49x^3 + 31.064x^2 + 17.252x$) was derived from the standard curve and used to calculate Adiponectin concentrations in samples based on their Average Absorbance values.

Experimental Notes: None

Images for Validation report #029593



Validation image no. 1 for Adiponectin (ADIPOQ) ELISA Kit (ABIN364986)

Figure 1: Graph of corrected OD450 nm plotted for standard curve samples.

Type	Sample, ng/ml	Readings at 450 nm			Avg Reading	Corrected OD ₄₅₀	SD	Calculated conc ng/ml
		1	2	3				
Standards ng/ml	100.00	2.241	2.030	2.136	2.136	2.070	0.086	100.00
	50.00	1.297	1.483	1.271	1.350	1.284	0.094	49.94
	25.00	0.800	0.899	0.782	0.827	0.781	0.051	25.27
	12.50	0.486	0.535	0.464	0.495	0.429	0.030	11.96
	6.25	0.372	0.338	0.282	0.331	0.265	0.037	6.46
	3.13	0.195	0.222	0.189	0.202	0.136	0.014	2.88
	1.56	0.180	0.175	0.167	0.174	0.108	0.005	2.21
	0.00	0.063	0.069	0.067	0.066	0.000	0.002	0.01
Spike Controls	6.25 ng/ml	0.333	0.345	0.341	0.340	0.274	0.005	6.73
	0 ng/ml	0.062	0.066	0.070	0.066	0.000	0.003	0.00
Samples	Serum, Human	3.855	3.891	3.944	3.897	3.831	0.037	503.83
	Serum, Human 1:1k diluted	0.421	0.400	0.433	0.415	0.352	0.014	9.27
	Serum, Goat	0.060	0.055	0.058	0.054	-0.008	0.002	-0.54

Adiponectin conc in Human Serum (+ve Control) : 9.27 ng/ml ± 1000 (dilution factor) = 9.27 µg/ml
 Adiponectin conc in undiluted human sera (503.83 ng/ml) is out-of-range of STD curve.
 Adiponectin conc in Goat Serum (-ve Control) : -0.14 ng/ml

Validation image no. 2 for Adiponectin (ADIPOQ) ELISA Kit (ABIN364986)

Table 1: ELISA. Adiponectin is present in human serum and undetectable in goat serum. Spike controls indicate no interference in absorbance readings from the diluent used to prepare standards and sera samples. Absorbance readings (OD450 nm) are shown for standard curve, spike controls and unknown samples. Value for Avg Reading is derived from the average reading of three samples. Avg Reading for "0" amount of Standard was subtracted from all

Avg Readings to yield "Corrected OD450 nm" values for standards, spike controls and unknown samples. Standard deviation is included for all samples. Standard curve was generated by regression analysis with four-parameter logistic. An equation ($y = 4.2198x^4 - 16.49x^3 + 31.064x^2 + 17.252x$) was derived from the standard curve and used to calculate Adiponectin concentrations shown in Table 1.