

Datasheet for ABIN366323

Insulin ELISA Kit

1 Image 26 Publications



Go to Product page

\sim				
	Ive	r\/		٨
\cup	' V C	ı vı	\Box	٧V

Quantity:	96 tests	
Target:	Insulin (INS)	
Reactivity:	Mouse	
Method Type:	Sandwich ELISA	
Detection Range:	15.6-1000 nIU/mL	
Minimum Detection Limit:	15.6 nIU/mL	
Application:	ELISA	
Product Details		
Purpose:	For the quantitative determination of mouse insulin (INS) concentrations in serum, plasma, cell culture supernates, tissue homogenates, cell lysates.	
Sample Type:	Serum, Plasma, Cell Culture Supernatant, Tissue Homogenate, Cell Lysate	
Analytical Method:	Quantitative	
Detection Method:	Colorimetric	
Specificity:	This assay has high sensitivity and excellent specificity for detection of mouse INS.	
Cross-Reactivity (Details):	Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between the target antigen and all analogues for other species. Therefore, cross reaction may still exist.	
Sensitivity:	3.9 nIU/mL	
Components:	Assay plate (12 × 8 coated Microwells)	

- · Standard (freeze dried)
- Biotin-antibody (100 × concentrate)
- HRP-avidin (100 × concentrate)
- · Biotin-antibody Diluent
- · HRP-avidin Diluent
- · Sample Diluent
- Wash Buffer (25 × concentrate)
- · TMB Substrate
- Stop Solution
- Adhesive Strip (for 96 wells)
- · Instruction manual

Target Details

Target:	Insulin (INS)
Abstract:	INS Products
Pathways:	NF-kappaB Signaling, RTK Signaling, Positive Regulation of Peptide Hormone Secretion, Peptide Hormone Metabolism, Hormone Activity, Carbohydrate Homeostasis, ER-Nucleus Signaling, Regulation of Carbohydrate Metabolic Process, Feeding Behaviour, Autophagy, Negative Regulation of intrinsic apoptotic Signaling, Brown Fat Cell Differentiation, Positive Regulation of fat Cell Differentiation

Application Details

Application Notes:

- The supplier is only responsible for the kit itself, but not for the samples consumed during the assay. The user should calculate the possible amount of the samples used in the whole test.
 Please reserve sufficient samples in advance.
- Samples to be used within 5 days may be stored at 2-8°C, otherwise samples must be stored at -20°C (≤ 1 month) or -80°C (≤ 2 months) to avoid loss of bioactivity and contamination.
- · Grossly hemolyzed samples are not suitable for use in this assay.
- If the samples are not indicated in the manual, a preliminary experiment to determine the validity of the kit is necessary.
- Please predict the concentration before assaying. If values for these are not within the range
 of the standard curve, users must determine the optimal sample dilutions for their particular
 experiments.
- Tissue or cell extraction samples prepared by chemical lysis buffer may cause unexpected ELISA results due to the impacts of certain chemicals.
- Owing to the possibility of mismatching between antigens from another resource and antibodies used in this supplier's kits (e.g., antibody targets conformational epitope rather than linear epitope), some native or recombinant proteins from other manufacturers may not

be recognized by this supplier's products.

- Influenced by factors including cell viability, cell number and cell sampling time, samples
 from cell culture supernatant may not be recognized by the kit.
- Fresh samples without long time storage are recommended for the test. Otherwise, protein degradation and denaturalization may occur in those samples and finally lead to wrong results.

Comment:

Detection wavelength: 450 nm

Information on standard material:

Depending on the antigen to be detected, standards can be either native or recombinant protein. The recombinant proteins are being expressed in CHO cells in most cases. Please inquire for more information. The formulation of auxiliary material in the standard is considered proprietary information, however it does not contain any poisonous substance. Proclin 300 (1:3000) is used as preservative.

Information on reagents:

In most cases the stop solution provided is 1 N H2SO4. The formulation of wash solution is proprietary information. None of the components contain (sodium) azide, thimerosal, 2-mercaptoethanol (2-ME) or any other poisonous materials. For the sandwich method kits, the sample diluent, antibody diluent, enzyme diluent and standard all contain BSA.

Information on antibodies:

The antibodies provided in different kits vary in regards to clonality and host. Some antibodies are affinity purified, some are Protein A

Sample Volume:

 $100~\mu L$

Assay Time:

1 - 4.5 h

Plate:

Pre-coated

Protocol:

This assay employs the quantitative sandwich enzyme immunoassay technique. Antibody specific for INS has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any INS present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for INS is added to the wells. After washing, avidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of INS bound in the initial step. The color development is stopped and the intensity of the color is measured.

Application Details

, ipplication because		
Assay Precision:	Intra-assay precision (precision within an assay): Three samples of known concentration were	
	tested twenty times on one plate to assess precision.	
	Inter-assay precision (precision between assays): Three samples of known concentration were	
	tested in twenty assays to assess precision.	
	Intra-assay: CV% less than 8%	
	Inter-assay: CV% less than 10%	
Restrictions:	For Research Use only	
Handling		
Precaution of Use:	The Stop Solution provided with this kit is an acid solution. Wear eye, hand, face and clothing	
	protection when using this material.	
Handling Advice:	The kit should not be used beyond the expiration date on the kit label.	
	 Do not mix or substitute reagents with those from other lots or sources. 	
	• If samples generate values higher than the highest standard, dilute the samples with Sample	
	Diluent and repeat the assay.	
	Any variation in Sample Diluent, operator, pipetting technique, washing technique, incubation	
	time/temperature and kit age can cause variation in binding.	
	This assay is designed to eliminate interference by soluble receptors, binding proteins and	
	other factors present in biological samples. Until all factors have been tested in the Immunoassay, the possibility of interference cannot be excluded.	
	infinditioassay, the possibility of interference cannot be excluded.	
Storage:	4 °C/-20 °C	
Storage Comment:	For unopened kit: All the reagents should be kept according to the labels on vials.	
Expiry Date:	6 months	
Publications		
Product cited in:	Li, Zhou, Chen, Zhang, Wang: "Kukoamine A attenuates insulin resistance and fatty liver through	
	downregulation of Srebp-1c." in: Biomedicine & pharmacotherapy, Vol. 89, pp. 536-543, (2018)	
	(PubMed).	
	Zeng, He, Jia, Hao: "Lycopene Improves Insulin Sensitivity through Inhibition of STAT3/Srebp-	
	2. 5, -1, 1.2, 1.2, 1.2, 1.2	

Zeng, He, Jia, Hao: "Lycopene Improves Insulin Sensitivity through Inhibition of STAT3/Srebp-1c-Mediated Lipid Accumulation and Inflammation in Mice fed a High-Fat Diet." in:

Experimental and clinical endocrinology & diabetes: official journal, German Society of Endocrinology [and] German Diabetes Association, Vol. 125, Issue 9, pp. 610-617, (2018) (PubMed).


Zhuang, Shou, Lu, Wang, Qiu, Wang, He, Chen, Jiao, Zhang: "Arachidonic acid sex-dependently affects obesity through linking gut microbiota-driven inflammation to hypothalamus-adipose-liver axis." in: **Biochimica et biophysica acta. Molecular basis of disease**, Vol. 1863, Issue 11, pp. 2715-2726, (2018) (PubMed).

Skurikhin, Pakhomova, Pershina, Ermolaeva, Krupin, Ermakova, Pan, Kudryashova, Rybalkina, Pavlovskaya, Litvyakov, Goldberg, Dygai: "Regenerative Potential of Spermatogonial Stem Cells, Endothelial Progenitor Cells, and Epithelial Progenitor Cells of C57Bl/6 Male Mice with Metabolic Disorders." in: **Bulletin of experimental biology and medicine**, Vol. 163, Issue 2, pp. 239-244, (2018) (PubMed).

Li, Wang, Xu, Luo, Luo, Hao, Cheng, Fang, Wang, Zhang, Chen: "Berberine Improves Diabetic Encephalopathy Through the SIRT1/ER Stress Pathway in db/db Mice." in: **Rejuvenation research**, Vol. 21, Issue 3, pp. 200-209, (2018) (PubMed).

There are more publications referencing this product on: Product page

Images



ELISA

Image 1. Typical standard curve