.-online.com antibodies

Datasheet for ABIN422478 **PNPT1 ELISA Kit**



Overview

Quantity:	96 tests
Target:	PNPT1
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.15 ng/mL - 10 ng/mL
Minimum Detection Limit:	0.15 ng/mL
Application:	ELISA

Product Details

Sample Type:	Plasma, Serum
Detection Method:	Colorimetric
Sensitivity:	0.054 ng/mL
Components:	Pre-coated, ready to use 96-well strip plate, flat buttom
	Plate sealer for 96 wells
	Reference Standard
	Standard Diluent
	Detection Reagent A
	Detection Reagent B
	Assay Diluent A
	Assay Diluent B
	Reagent Diluent (if Detection Reagent is lyophilized)
	TMB Substrate
	Stop Solution

- Wash Buffer (30 x concentrate)
- Instruction manual

Target Details

Target:	PNPT1
Alternative Name:	Polyribonucleotide Nucleotidyltransferase 1 (PNPT1 Products)
Application Details	
Application Notes:	 Limited by the current condition and scientific technology, we cannot completely conduct the comprehensive identification and analysis on the raw material provided by suppliers. So there might be some qualitative and technical risks to use the kit. The final experimental results will be closely related to validity of the products, operation skills of the end users and the experimental environments. Please make sure that sufficient samples are available. Kits from different batches may be a little different in detection range, sensitivity and color developing time. Do not mix or substitute reagents from one kit lot to another. Use only the reagents supplied by manufacturer. Protect all reagents from strong light during storage and incubation. All the bottle caps of reagents should be covered tightly to prevent the evaporation and contamination of microorganism. There may be some foggy substance in the wells when the plate is opened at the first time. If will not have any effect on the final assay results. Do not remove microtiter plate from the storage bag until needed. Wrong operations during the reagents preparation and loading, as well as incorrect parameter setting for the plate reader may lead to incorrect results. A microplate plate readee with a bandwidth of 10m or less and an optical density range of 0-3 0.D. or greater at 450 ± 10nm wavelength is acceptable for use in absorbance measurement. Please read the instruction carefully and adjust the instrument prior to the experiment. Even the same operator might get different results in two separate experiments. In order to get better reproducible results, the operation of every step in the assay should be controlled. Furthermore, a preliminary experiment before assay for each batch is recommended. Each kit has been strictly passed Q.C test. However, results from end users might be inconsistent with our in-house data due to some unexpected transportation conditions or different
Comment:	Information on standard material:

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/4 | Product datasheet for ABIN422478 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

Application Details	
	The standard might be recombinant protein or natural protein, that will depend on the specific
	kit. Moreover, the expression system is E.coli or yeast or mammal cell. There is 0.05% proclin
	300 in the standard as preservative.
	Information on reagents:
	The stop solution used in the kit is sulfuric acid with concentration of 1 mol/L. And the wash
	solution is TBS. The standard diluent contains 0.02 % sodium azide, assay diluent A and assay
	diluent B contain 0.01% sodium azide. Some kits can contain is BSA in them.
	Information on antibodies:
	The provided antibodies and their host vary in different kits.
Sample Volume:	100 µL
Plate:	Pre-coated
Plate: Reagent Preparation:	 Bring all kit components and samples to room temperature (18-25 °C) before use. If the kit will not be used up in one time, please only take out strips and reagents for present experiment, and leave the remaining strips and reagents in required condition. Standard - Reconstitute the Standard with 1.0 mL of Standard Diluent, keep for 10 minutes a room temperature, shake gently (not to foam). The concentration of the standard in the stoce solution is 10 ng/mL. Prepare 7 tubes containing 0.5 mL Standard Diluent and produce a double dilution series. Mix each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 10 ng/mL, 5 ng/mL, 2.5 ng/mL, 1.25 ng/mL, 0.625 ng/mL, 0.312 ng/mL, 0.156 ng/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 0 ng/mL. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x). TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.
	Note:
	1. Making serial dilution in the wells directly is not permitted.
	2. Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.
	3. Please carefully reconstitute Standards or working Detection Reagent A and B according to
	the instruction, and avoid foaming and mix gently until the crystals are completely dissolved.

Application Details

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 3/4 | Product datasheet for ABIN422478 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

Expiry Date:	days at 37°C equaling 6 months at 4°C. 6 months
Storage Comment:	 For unopened kit: All the reagents should be kept according to the labels on vials. The Standard, Detection Reagent A, Detection Reagent B and the 96-well strip plate should be stored at -20°C upon receipt while the others should be at 4°C. For opened kit: When the kit is opened, the remaining reagents still need to be stored according to the above storage condition. Besides, please return the unused wells to the foil pouch containing the desiccant pack, and reseal along entire edge of zip-seal. Note: It is highly recommended to use the remaining reagents within 1 month provided this is within the expiration date of the kit. For ELISA kit, 1 day storage at 37°C can be considered as 2 months at 4°C, which means 3
Storage:	4 °C/-20 °C
Handling Advice:	To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Precaution of Use:	The Stop Solution suggested for use with this kit is an acid solution. Wear eye, hand, face, and clothing protection when using this material.
Handling	
Restrictions:	For Research Use only
	 Intra-assay: CV less than 10 % Inter-assay: CV less than 12 %
	CV (%) = SD/mean X 100
	levels of the target antigen were tested on three different plates, eight replicates in each plate.
	levels of the target antigen were tested twenty times on one plate, respectively. Inter-assay Precision (precision between assays): Three samples with low, medium and high
Assay Precision:	Intra-assay Precision (precision within an assay): Three samples with low, medium and high
	 4. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once. 5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature and mix gently until the crystals are completely dissolved. 6. Contaminated water or container for reagent preparation will influence the detection result.
	To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.

Expiry Date:

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 4/4 | Product datasheet for ABIN422478 | 09/12/2023 | Copyright antibodies-online. All rights reserved.