## .-online.com antibodies

# Datasheet for ABIN454288 Acrosin ELISA Kit



Overview

Quantity:	96 tests
Target:	Acrosin (ACR)
Reactivity:	Human
Method Type:	Competition ELISA
Detection Range:	6.25-400 U/L
Minimum Detection Limit:	6.25 U/L
Application:	ELISA

#### Product Details

Purpose:	This immunoassay kit allows for the use in vitro quantitative determination of human anti- acrosin antibodies, AR Ab concentrations in cell culture supernates, serum, plasma and other biological fluids.
Sample Type:	Cell Culture Supernatant, Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay recognizes recombinant and natural human AR Ab.
Cross-Reactivity (Details):	No significant cross-reactivity or interference was observed.
Sensitivity:	< 3.12 U/L
	The sensitivity of this assay, or Lower Limit of Detection (LLD) was defined as the lowest
	detectable concentration that could be differentiated from zero.

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/5 | Product datasheet for ABIN454288 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

Product Details	
Characteristics:	Homo sapiens,Human,Acrosin,ACR,ACRS,3.4.21.10
Components:	Reagent (Quantity): Assay plate (1), Standard (2), Sample Diluent (1x20ml), Assay Diluent A (1x10ml), Assay DiluentB 1 x 10ml Detection Reagent A (1x120µl), Detection Reagent B
	(1x120µI), Wash Buffer(25 x concentrate) (1x30mI), Substrate (1x10mI), Stop Solution (1x10mI),
	Plate sealer for 96 wells (1x5)

### Target Details

Target:	Acrosin (ACR)
Alternative Name:	ACR (ACR Products)
Pathways:	cAMP Metabolic Process

### Application Details

Sample Volume:	100 µL
Plate:	Pre-coated
Protocol:	The microtiter plate provided in this kit has been pre-coated with AR antigen. Standards or
	samples are then added to the appropriate microtiter plate wells with a biotin-conjugated AR
	antigen and Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate
	well and incubated. Then a TMB (3,3'5, 5' tetramethyl-benzidine) substrate solution is added to
	each well. Only those wells that contain AR Ab, biotin-conjugated antigen and enzyme-
	conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by
	the addition of a sulphuric acid solution and the color change is measured
	spectrophotometrically at a wavelength of 450 nm $\pm$ 2 nm. The concentration of AR Ab in the
	samples is then determined by comparing the O.D. of the samples to the standard curve.
Reagent Preparation:	Bring all reagents to room temperature before use. Wash Buffer - If crystals have formed in the
	concentrate, warm to room temperature and mix gently until the crystals have completely
	dissolved. Dilute 30 mL of Wash Buffer Concentrate into deionized or distilled water to prepare
	750 mL of Wash Buffer. Standard - Reconstitute the Standard with 1.0 mL of Sample Diluent.
	This reconstitution produces a stock solution of 400 U/L. Allow the standard to sit for a
	minimum of 15 minutes with gentle agitation prior to making serial dilutions. The undiluted
	standard serves as the high standard (400 U/L). The Sample Diluent serves as the zero
	standard (0 U/L). Detection Reagent A and B - Dilute to the working concentration specified on
	the vial label using Assay Diluent A and B (1:100), respectively.

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/5 | Product datasheet for ABIN454288 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

#### Application Details

Sample Collection: Serum - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before 2 centrifugation for 15 minutes at approximately 1000 x g. Remove serum and assay immediately or aliquot and store samples at -20 °C or -80 °C. Plasma - Collect plasma using EDTA or heparin as an anticoagulant. Centrifuge samples for 15 minutes at 1000 x g at 2 - 8 °C within 30 minutes of collection. Store samples at -20 °C or -80 °C. Avoid repeated freeze-thaw cycles. Cell culture supernates and other biological fluids - Remove particulates by centrifugation and assay immediately or aliquot and store samples at -20 °C or -80 °C. Avoid repeated freeze-thaw cycles. Note: Serum, plasma, and cell culture supernatant samples to be used within 7 days may be stored at 2-8C, otherwise samples must stored at -20 °C (< 3 months) or -80 °C (< 6 months) to avoid loss of bioactivity and contamination. Avoid freeze-thaw cycles. When performing the assay slowly bring samples to room temperature. It is recommended that all samples be assayed in duplicate.

Assay Procedure:

Allow all reagents to reach room temperature. All the reagents should be mixed thoroughly by gently swirling before pipetting. Avoid foaming. Arrange and label required number of strips. Prepare all reagents, working standards and samples as directed in the previous sections. 3 1. Add 100 uL of Standard, Blank, or Sample per well. Cover with the Plate sealer. Incubate for 2 hours at 37 °C.

2. Remove the liquid of each well, don't wash.

3. Add 100 uL of Detection Reagent A working solution to each well. Cover with the Plate sealer. Incubate for 1 hour at 37°C. Detection Reagent A working solution may appear cloudy. Warm to room temperature and mix gently until solution appears uniform.

4. Aspirate each well and wash, repeating the process three times for a total of three washes. Wash by filling each well with Wash Buffer (350 uL) using a squirt bottle, multi-channel pipette, manifold dispenser or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.

5. Add 100 uL of Detection Reagent B working solution to each well. Cover with a new Plate sealer. Incubate for 1 hours at 37 °C.

6. Repeat the aspiration/wash as in step

4. 7. Add 90 uL of Substrate Solution to each well. Cover with a new Plate sealer. Incubate within 30 minutes at 37°C. Protect from light.

8. Add 50 uL of Stop Solution to each well. If color change does not appear uniform, gently tap the plate to ensure thorough mixing.

9. Determine the optical density of each well at once, using a microplate reader set to 450 nm. Important Note:

	1. Please carefully reconstitute Standards or working Detection Reagent A and B according to
	the instruction, and avoid foaming and mix gently until the crystals have completely dissolved.
	The reconstituted Standards can be used only once.
	2. The wash procedure is critical. Insufficient washing will result in poor precision and falsely
	elevated absorbance readings.
	3. It is recommended that no more than 32 wells be used for each assay run if manual pipetting
	is used since pipetting of all standards, specimens and controls should be completed within 5
	minutes. A full plate of 96 wells may be used if automated pipetting is available.
	4. Duplication of all standards and specimens, although not required, is recommended.
	5. When mixing or reconstituting protein solutions, always avoid foaming. 4
	6. To avoid cross-contamination, change pipette tips between additions of each standard level,
	between sample additions, and between reagent additions. Also, use separate reservoirs for each reagent.
	7. To ensure accurate results, proper adhesion of plate sealers during incubation steps is necessary.
	8. Do not substitute reagents from one kit lot to another. Use only the reagents supplied by
	manufacturer.
Calculation of Results:	Average the duplicate readings for each standard, control, and sample and subtract the average
	zero standard optical density. Create a standard curve by reducing the data using computer
	software capable of generating a four parameter logistic (4-PL) curve-fit. As an alternative,
	construct a standard curve by plotting the mean absorbance for each standard on the y-axis
	against the concentration on the x-axis and draw a best fit curve through the points on the
	graph. The data may be linearized by plotting the log of the AR Ab concentrations versus the log
	of the O.D. and the best fit line can be determined by regression analysis. This procedure will
	produce an adequate but less precise fit of the data. If samples have been diluted, the
	concentration read from the standard curve must be multiplied by the dilution factor.
Restrictions:	For Research Use only
Handling	
Handling Advice:	1. The kit should not be used beyond the expiration date on the kit label.
	2. Do not mix or substitute reagents with those from other lots or sources.
	3. If samples generate values higher than the highest standard, further dilute the samples with
	the Assay Diluent and repeat the assay. Any variation in standard diluent, operator, pipetting
	technique, washing technique,incubation time or temperature, and kit age can cause variation ir
	binding.

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 4/5 | Product datasheet for ABIN454288 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

	4. This assay is designed to eliminate interference by soluble receptors, ligands, binding
	proteins, and other factors present in biological samples. Until all factors have been tested in
	the Immunoassay, the possibility of interference cannot be excluded.
Storage:	4 °C/-20 °C
Storage Comment:	The Standard, Detection Reagent A, Detection Reagent B and the 96-well strip plate should be
	stored at -20 °C upon being received. The other reagents can be stored at 4 °C.